



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87380

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/10/2017	18:15
2	CAL2	1	Water			7.00	05/10/2017	18:16
3	CAL3	1	Water			10.04	05/10/2017	18:17
4	ICV	1	Water			7.01	05/10/2017	18:19
5	CCV1	1	Water			2.02	05/10/2017	18:20
6	I3084-01	1	Solid	20.02	20	6.48	05/10/2017	18:21
7	I3084-01DUP	1	Solid	20.03	20	6.49	05/10/2017	18:22
8	I3087-01	1	Solid	20.04	20	6.11	05/10/2017	18:23
9	CCV2	1	Water			12.02	05/10/2017	18:25

LB 87380

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 13084

WorkList ID : 98321

Date : 5/10/2017 6:02:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3084-01	pH	Cool 4 deg C	TULL02	G32	FK-PS-01-001	05/10/2017	9045C
	Solid	I3087-01	pH	Cool 4 deg C	TULL02	G32	FK-PS-01_001	05/10/2017	9045C

Date/Time 05-10-17 6:02 PM
Received by: JP
Relinquished by: JP

Date/Time 05-10-17 7:10 PM
Received by: JP
Relinquished by: JP