



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87388

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/11/2017	15:30
2	CAL2	1	Water			7.01	05/11/2017	15:31
3	CAL3	1	Water			10.03	05/11/2017	15:32
4	ICV	1	Water			6.99	05/11/2017	15:35
5	CCV1	1	Water			2.02	05/11/2017	15:36
6	I3069-02	1	Solid	20.02	20	10.02	05/11/2017	15:37
7	I3092-07	1	Solid	20.04	20	6.85	05/11/2017	15:39
8	I3092-07DUP	1	Solid	20.01	20	6.86	05/11/2017	15:40
9	I3092-08	1	Solid	20.02	20	6.55	05/11/2017	15:41
10	I3092-09	1	Solid	20.05	20	10.58	05/11/2017	15:42
11	CCV2	1	Water			12.01	05/11/2017	15:45

WORKLIST(Hardcopy Internal Chain)

1787388

WorkList Name : CORROSIVITY I3069

WorkList ID : 98318

Date : 5/10/2017 3:18:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3069-02	Corrosivity	Cool 4 deg C	VIAS01	G42	GAC-2	05/09/2017	9045C
	Solid	I3092-07	Corrosivity	Cool 4 deg C	LIRO01	G62	SB-04-COMP	05/09/2017	9045C
	Solid	I3092-08	Corrosivity	Cool 4 deg C	LIRO01	G62	SB-05-COMP	05/09/2017	9045C
	Solid	I3092-09	Corrosivity	Cool 4 deg C	LIRO01	G62	SB-07-COMP	05/09/2017	9045C

Date/Time 05-10-17 3:18 PM
 Received by: JP
 Relinquished by: JP

Date/Time 05-10-17 6:30 PM
 Received by: JP
 Relinquished by: JP