


7470A

LB87397
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36029



A= 0.0000e+000

B= 2.7647e-004

C= -7.1391e-003

Rho= 0.9999353

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	83	0.000	83				
0.2	0.200	0.203	0.003	761	0.0 %	761				
2.5	2.500	2.444	-0.056	8866	0.0 %	8866				
5.0	5.000	5.013	0.013	18158	0.0 %	18158				
7.5	7.500	7.570	0.070	27406	0.0 %	27406				
10.0	10.000	9.954	-0.046	36029	0.0 %	36029				

LB87397 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 May 2017 12:13:31

Sample ID	Extended ID	μ Abs.	Conc.	Std. Cond	Method	Units	Date	Type	Type
0.0 - 1	50	83	-	0.0007470A		PPB	11 May 2017 12:21:02	S	Std
0.2 - 1	50.2	761	-	0.20007470A		PPB	11 May 2017 12:23:10	S	Std
2.5 - 1	52.5	8866	-	2.50007470A		PPB	11 May 2017 12:25:19	S	Std
5.0 - 1	55	18158	-	5.00007470A		PPB	11 May 2017 12:27:27	S	Std
7.5 - 1	57.5	27406	-	7.50007470A		PPB	11 May 2017 12:29:35	S	Std
10.0 - 1	510	36029	-	10.00007470A		PPB	11 May 2017 12:34:11	S	Std
ICV52 - 1	ICV52	13823	3.8146	-7470A		PPB	11 May 2017 12:36:38	U	SMPL
ICB52 - 1	ICB52	-6	-0.0088	-7470A		PPB	11 May 2017 12:38:44	U	SMPL
CCV83 - 1	CCV83	18304	5.0534	-7470A		PPB	11 May 2017 12:43:24	U	SMPL
CCB83 - 1	CCB83	-25	-0.0141	-7470A		PPB	11 May 2017 12:45:30	U	SMPL
CRA - 1	CRA	745	0.1988	-7470A		PPB	11 May 2017 12:47:36	U	SMPL
HighStd - 1	HighStd	36438	10.0670	-7470A		PPB	11 May 2017 12:49:43	U	SMPL
ChkStd - 1	ChkStd	25248	6.9733	-7470A		PPB	11 May 2017 12:51:50	U	SMPL
PB98855BL - 1	PBW	-76	-0.0282	-7470A		PPB	11 May 2017 12:53:58	U	SMPL
PB98855BS - 1	LCSW	13939	3.8466	-7470A		PPB	11 May 2017 12:56:05	U	SMPL
PB98855TB - 1	PB98827TB	-39	-0.0179	-7470A		PPB	11 May 2017 12:58:13	U	SMPL
I3046-09 - 1	SB-11-COMP	13	-0.0035	-7470A		PPB	11 May 2017 13:00:20	U	SMPL
I3046-10 - 1	SB-12-COMP	50	0.0067	-7470A		PPB	11 May 2017 13:02:29	U	SMPL
I3046-11 - 1	SB-13-COMP	43	0.0047	-7470A		PPB	11 May 2017 13:04:37	U	SMPL
I3059-03 - 1	JC-SWITCH-DRUM-SPILL(COMP)	48	0.0061	-7470A		PPB	11 May 2017 13:06:44	U	SMPL
CCV84 - 1	CCV84	18223	5.0310	-7470A		PPB	11 May 2017 13:08:51	U	SMPL
CCB84 - 1	CCB84	-22	-0.0132	-7470A		PPB	11 May 2017 13:10:58	U	SMPL
I3062-16 - 1	SB-10-COMP	32	0.0017	-7470A		PPB	11 May 2017 13:13:05	U	SMPL
I3062-17 - 1	SB-09-COMP	35	0.0025	-7470A		PPB	11 May 2017 13:15:11	U	SMPL
I3062-18 - 1	SB-08-COMP	37	0.0031	-7470A		PPB	11 May 2017 13:17:18	U	SMPL
I3062-19 - 1	SB-14-COMP	35	0.0025	-7470A		PPB	11 May 2017 13:19:25	U	SMPL
I3062-20 - 1	SB-06-COMP	31	0.0014	-7470A		PPB	11 May 2017 13:21:33	U	SMPL
I3062-21 - 1	SB-02-COMP	63	0.0103	-7470A		PPB	11 May 2017 13:23:41	U	SMPL
I3062-22 - 1	SB-03-COMP	53	0.0075	-7470A		PPB	11 May 2017 13:25:49	U	SMPL
I3062-22DUP - 1	SB-03-COMP DUP	49	0.0064	-7470A		PPB	11 May 2017 13:27:57	U	SMPL
I3062-22MS - 1	SB-03-COMPMS	13734	3.7899	-7470A		PPB	11 May 2017 13:30:06	U	SMPL
I3062-22MSD - 1	SB-03-COMPMSD	13373	3.6901	-7470A		PPB	11 May 2017 13:32:13	U	SMPL
CCV85 - 1	CCV85	18215	5.0288	-7470A		PPB	11 May 2017 13:34:20	U	SMPL
CCB85 - 1	CCB85	-23	-0.0135	-7470A		PPB	11 May 2017 13:36:26	U	SMPL
PB98856BL - 1	PBW	24	-0.0005	-7470A		PPB	11 May 2017 13:38:33	U	SMPL
PB98856BS - 1	LCSW	13388	3.6943	-7470A		PPB	11 May 2017 13:40:40	U	SMPL
I3052-01 - 1	MW-1-20170508	-38	-0.0176	-7470A		PPB	11 May 2017 13:42:47	U	SMPL
I3052-01DUP - 1	MW-1-20170508DUP	28	0.0006	-7470A		PPB	11 May 2017 13:44:53	U	SMPL
I3052-01MS - 1	MW-1-20170508MS	14451	3.9882	-7470A		PPB	11 May 2017 13:47:01	U	SMPL
I3052-01MSD - 1	MW-1-20170508MSD	14188	3.9155	-7470A		PPB	11 May 2017 13:49:08	U	SMPL
I3052-02 - 1	MW-2-20170508	98	0.0200	-7470A		PPB	11 May 2017 13:51:16	U	SMPL
I3052-04 - 1	MW-1-20170508	-17	-0.0118	-7470A		PPB	11 May 2017 13:53:24	U	SMPL
I3052-05 - 1	MW-2-20170508	18	-0.0022	-7470A		PPB	11 May 2017 13:55:33	U	SMPL
I3061-05 - 1	VL-WWTP-002	44	0.0050	-7470A		PPB	11 May 2017 13:57:39	U	SMPL
CCV86 - 1	CCV86	17694	4.8848	-7470A		PPB	11 May 2017 13:59:46	U	SMPL
CCB86 - 1	CCB86	-37	-0.0174	-7470A		PPB	11 May 2017 14:01:53	U	SMPL
I3062-22LX5 - 1		48	0.0061	-7470A		PPB	11 May 2017 14:04:00	U	SMPL
I3062-22A - 1		14391	3.9716	-7470A		PPB	11 May 2017 14:06:08	U	SMPL
I3052-01LX5 - 1		-13	-0.0107	-7470A		PPB	11 May 2017 14:08:15	U	SMPL
I3052-01A - 1		14819	4.0899	-7470A		PPB	11 May 2017 14:10:22	U	SMPL
CCV87 - 1	CCV87	18301	5.0526	-7470A		PPB	11 May 2017 14:12:30	U	SMPL
CCB87 - 1	CCB87	-43	-0.0190	-7470A		PPB	11 May 2017 14:14:37	U	SMPL