


ISM02.4

LB 87401
INSTRUMENT ID: CY1

Linear

 μ Abs.:

45870



A= 0.0000e+000

B= 2.1679e-004

C= -1.8938e-002

Rho= 0.9998884

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.010	-0.010	43	0.000	43					-
0.2	0.200	0.202	0.002	1021	0.0 %	1021					1
2.5	2.500	2.483	-0.017	11540	0.0 %	11540					-1
5.0	5.000	4.989	-0.011	23099	0.0 %	23099					0
7.5	7.500	7.610	0.110	35191	0.0 %	35191					1
10.0	10.000	9.925	-0.075	45870	0.0 %	45870					-1

203
5/10/17

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
0 S0		43	-	0	ISM02.4	PPB	5/11/2017 14:41	Std
0.2 S0.2		1021	-	0.2	ISM02.4	PPB	5/11/2017 14:43	Std
2.5 S2.5		11540	-	2.5	ISM02.4	PPB	5/11/2017 14:46	Std
5 S5		23099	-	5	ISM02.4	PPB	5/11/2017 14:48	Std
7.5 S7.5		35191	-	7.5	ISM02.4	PPB	5/11/2017 14:50	Std
10 S10		45870	-	10	ISM02.4	PPB	5/11/2017 14:52	Std
ICV	ICV	17933	3.8688 -		ISM02.4	PPB	5/11/2017 14:55	SMPL
ICB	ICB	-16	-0.0224 -		ISM02.4	PPB	5/11/2017 14:58	SMPL
CCV088	CCV088	22929	4.9519 -		ISM02.4	PPB	5/11/2017 15:00	SMPL
CCB088	CCB088	-19	-0.0231 -		ISM02.4	PPB	5/11/2017 15:02	SMPL
PB98860BL	PBS004	49	-0.0083 -		ISM02.4	PPB	5/11/2017 15:04	SMPL
I2962-01	MBDC41	11995	2.5815 -		ISM02.4	PPB	5/11/2017 15:07	SMPL
I2962-02	MBDC40	13709	2.9531 -		ISM02.4	PPB	5/11/2017 15:14	SMPL
I2962-03	MBDC40D	13623	2.9344 -		ISM02.4	PPB	5/11/2017 15:16	SMPL
I2962-04	MBDC40S	24351	5.2602 -		ISM02.4	PPB	5/11/2017 15:18	SMPL
I2962-05	MBDC39	925	0.1816 -		ISM02.4	PPB	5/11/2017 15:21	SMPL
I2962-06	MBDC43	1194	0.2399 -		ISM02.4	PPB	5/11/2017 15:23	SMPL
CCV089	CCV089	23353	5.0438 -		ISM02.4	PPB	5/11/2017 15:25	SMPL
CCB089	CCB089	-3	-0.0196 -		ISM02.4	PPB	5/11/2017 15:28	SMPL