

LB87406

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/11/2017 15:56

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.993	0.0	0.114	
ICB1	0.003	0.0	0.002	
CCV1	0.998	0.0	0.114	
CCB1	0.005	0.0	0.002	
PB98878BL	0.007	0.0	0.002	
PB98878BS	0.999	0.0	0.114	
I3062-15	0.002	0.0	0.002	
I3062-15DUP	0.008	0.0	0.002	
I3062-15MS	0.802	0.0	0.092	
I3062-15MSD	0.802	0.0	0.092	
CCV2	1.002	0.0	0.115	
CCB2	-0.000	0.0	0.002	

N	12
Mean	0.469
SD	0.4898
CV%	104.54

LB 87406

Aquakem v. 7.2AQ1

Results from time period:

Thu May 11 15:20:28 2017

AK

Thu May 11 15:52:06 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0535	mg/l	5/11/2017 15:20:28	
0.1PPM	A	Phenolics-	P	0.1056	mg/l	5/11/2017 15:20:29	
0.25PPM	A	Phenolics-	P	0.2529	mg/l	5/11/2017 15:20:30	
0.50PPM	A	Phenolics-	P	0.4966	mg/l	5/11/2017 15:20:31	
1.0PPM	A	Phenolics-	P	0.9827	mg/l	5/11/2017 15:20:32	
2.0PPM	A	Phenolics-	P	2.0088	mg/l	5/11/2017 15:20:33	
ICV1	S	Phenolics-	P	0.9934	mg/l	5/11/2017 15:49:18	
ICB1	S	Phenolics-	P	0.0027	mg/l	5/11/2017 15:49:19	
CCV1	S	Phenolics-	P	0.9982	mg/l	5/11/2017 15:49:20	
CCB1	S	Phenolics-	P	0.0051	mg/l	5/11/2017 15:49:21	
PB98878BL	S	Phenolics-	P	0.0075	mg/l	5/11/2017 15:49:22	
PB98878BS	S	Phenolics-	P	0.9993	mg/l	5/11/2017 15:49:23	
I3062-15	S	Phenolics-	P	0.0023	mg/l	5/11/2017 15:49:24	
I3062-15DUP	S	Phenolics-	P	0.0075	mg/l	5/11/2017 15:49:25	
I3062-15MS	S	Phenolics-	P	0.802	mg/l	5/11/2017 15:49:26	
I3062-15MSD	S	Phenolics-	P	0.8023	mg/l	5/11/2017 15:49:27	
CCV2	S	Phenolics-	P	1.0023	mg/l	5/11/2017 15:49:28	
CCB2	S	Phenolics-	P	-0.0001	mg/l	5/11/2017 15:49:29	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/11/2017 15:25


Test Phenolics-

Accepted 5/11/2017 15:25

Factor 8.866
Bias 0.002

Coeff. of det. 0.999843

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0535	0.0500	
2	Phenol-2	0.013	0.1056	0.1000	
3	Phenol-2	0.030	0.2529	0.2500	
4	Phenol-2	0.058	0.4966	0.5000	
5	Phenol-2	0.112	0.9827	1.0000	
6	Phenol-2	0.228	2.0088	2.0000	