



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB87408

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	05/11/2017	11:30
2	I3046-01		1	>100.0	>212.0	05/11/2017	11:55
3	I3046-01DUP		1	>100.0	>212.0	05/11/2017	12:25
4	I3046-02		1	>100.0	>212.0	05/11/2017	13:00
5	I3062-15		1	>100.0	>212.0	05/11/2017	13:23

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 87408

WorkList Name : FP51117

WorkList ID : 98366

Date : 5/11/2017 10:00AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3046-01	Flash Point	Cool 4 deg C	LIR001	F62	TWP-13	05/08/2017	1010A
	Water	I3046-02	Flash Point	Cool 4 deg C	LIR001	F62	TWP-11	05/08/2017	1010A
	Water	I3062-15	Flash Point	Cool 4 deg C	LIR001	G52	TWP-02	05/09/2017	1010A

Date/Time

5/11/17 10:10 AM

Received by:

KS

Relinquished by:

CS

Date/Time

5/11/17 13:30

Received by:

CS

Relinquished by:

KS