


ISM02.4

L387415
INSTRUMENT ID: CYA

Linear

 μ Abs.:

42520



A= 0.0000e+000

B= 2.3642e-004

C= 1.9527e-002

Rho= 0.9999194

Accept=Accepted

dope
r=1315 accept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.018	0.018	-6	0.000	-6				
0.2	0.200	0.216	0.016	833	0.0 %	833				
2.5	2.500	2.512	0.012	10544	0.0 %	10544				
5.0	5.000	4.950	-0.050	20856	0.0 %	20856				
7.5	7.500	7.431	-0.069	31349	0.0 %	31349				
10.0	10.000	10.072	0.072	42520	0.0 %	42520				

9010
8
0
-1
-1
1
203
5/14/17

LB87415
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-6	-		0 ISM02.4	PPB	5/12/2017 11:03	Std
	0.2 S0.2	833	-		0.2 ISM02.4	PPB	5/12/2017 11:05	Std
	2.5 S2.5	10544	-		2.5 ISM02.4	PPB	5/12/2017 11:08	Std
	5 S5	20856	-		5 ISM02.4	PPB	5/12/2017 11:10	Std
	7.5 S7.5	31349	-		7.5 ISM02.4	PPB	5/12/2017 11:12	Std
	10 S10	42520	-		10 ISM02.4	PPB	5/12/2017 11:20	Std
ICV	ICV	16703	3.9684 -		ISM02.4	PPB	5/12/2017 11:22	SMPL
ICB	ICB	-114	-0.0074 -		ISM02.4	PPB	5/12/2017 11:25	SMPL
CCV090	CCV090	20658	4.9034 -		ISM02.4	PPB	5/12/2017 11:27	SMPL
CCB090	CCB090	-136	-0.0126 -		ISM02.4	PPB	5/12/2017 11:29	SMPL
PB98880BL	PBW003	-38	0.0105 -		ISM02.4	PPB	5/12/2017 11:31	SMPL
I2993-01	MBDCP4	-46	0.0087 -		ISM02.4	PPB	5/12/2017 11:34	SMPL
I2993-02	MBDCP4D	-39	0.0103 -		ISM02.4	PPB	5/12/2017 11:36	SMPL
I2993-03	MBDCP4S	4190	1.0101 -		ISM02.4	PPB	5/12/2017 11:38	SMPL
I2993-04	MBDCP5	-63	0.0046 -		ISM02.4	PPB	5/12/2017 11:41	SMPL
I2993-05	MBDCP6	-73	0.0023 -		ISM02.4	PPB	5/12/2017 11:43	SMPL
I2993-06	MBDCT9	-94	-0.0027 -		ISM02.4	PPB	5/12/2017 11:45	SMPL
I2993-07	MBDCW2	-69	0.0032 -		ISM02.4	PPB	5/12/2017 11:47	SMPL
I2993-08	MBDCW3	-62	0.0049 -		ISM02.4	PPB	5/12/2017 11:50	SMPL
I2993-09	MBDCQ2	-39	0.0103 -		ISM02.4	PPB	5/12/2017 11:52	SMPL
I2993-10	MBDCQ3	-75	0.0018 -		ISM02.4	PPB	5/12/2017 11:54	SMPL
I2993-11	MBDCQ4	-58	0.0058 -		ISM02.4	PPB	5/12/2017 11:56	SMPL
I2993-13	MBDCS0	190	0.0644 -		ISM02.4	PPB	5/12/2017 11:59	SMPL
PB98881BL	PBW003	-23	0.0141 -		ISM02.4	PPB	5/12/2017 12:01	SMPL
I2994-01	MBDCE4	-35	0.0113 -		ISM02.4	PPB	5/12/2017 12:03	SMPL
I2994-02	MBDCE4D	-37	0.0108 -		ISM02.4	PPB	5/12/2017 12:06	SMPL
I2994-03	MBDCE4S	4197	1.0118 -		ISM02.4	PPB	5/12/2017 12:08	SMPL
I2994-04	MBDCE5	-27	0.0131 -		ISM02.4	PPB	5/12/2017 12:10	SMPL
I2994-05	MBDCE6	68	0.0356 -		ISM02.4	PPB	5/12/2017 12:12	SMPL
I2994-06	MBDCJ9	-86	-0.0008 -		ISM02.4	PPB	5/12/2017 12:15	SMPL
CCV091	CCV091	20938	4.9696 -		ISM02.4	PPB	5/12/2017 12:17	SMPL
CCB091	CCB091	-109	-0.0062 -		ISM02.4	PPB	5/12/2017 12:19	SMPL
I2994-07	MBDCK0	-78	0.0011 -		ISM02.4	PPB	5/12/2017 12:22	SMPL
I2994-08	MBDCK1	-25	0.0136 -		ISM02.4	PPB	5/12/2017 12:24	SMPL
I2994-09	MBDCK2	-39	0.0103 -		ISM02.4	PPB	5/12/2017 12:26	SMPL
I2994-10	MBDCK3	-29	0.0127 -		ISM02.4	PPB	5/12/2017 12:28	SMPL
I2994-11	MBDCF2	277	0.085 -		ISM02.4	PPB	5/12/2017 12:31	SMPL
I2994-12	MBDCF3	-49	0.0079 -		ISM02.4	PPB	5/12/2017 12:33	SMPL
I2994-13	MBDCF4	79	0.0382 -		ISM02.4	PPB	5/12/2017 12:35	SMPL
I2994-14	MBDCK4	-82	0.0001 -		ISM02.4	PPB	5/12/2017 12:37	SMPL
I2994-16	MBDCH0	1813	0.4481 -		ISM02.4	PPB	5/12/2017 12:40	SMPL
CCV092	CCV092	20640	4.8992 -		ISM02.4	PPB	5/12/2017 12:42	SMPL
CCB092	CCB092	-101	-0.0044 -		ISM02.4	PPB	5/12/2017 12:44	SMPL