


7471A

LB87426
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44627



A= 0.0000e+000

B= 2.2501e-004

C= 1.1200e-002

Rho= 0.9996009

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	-35	0.000	-35				
0.2	0.200	0.216	0.016	911	0.0 %	911				
2.5	2.500	2.426	-0.074	10733	0.0 %	10733				
5.0	5.000	5.169	0.169	22923	0.0 %	22923				
7.5	7.500	7.332	-0.168	32537	0.0 %	32537				
10.0	10.000	10.053	0.053	44627	0.0 %	44627				

LB87426 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 12 May 2017 14:42:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-35	-	0.0000	7471A	PPB	12 May 2017 14:48:37	S	Std
0.2 - 1	50.2	911	-	0.2000	7471A	PPB	12 May 2017 14:50:54	S	Std
2.5 - 1	52.5	10733	-	2.5000	7471A	PPB	12 May 2017 14:53:12	S	Std
5.0 - 1	55	22923	-	5.0000	7471A	PPB	12 May 2017 14:55:29	S	Std
7.5 - 1	57.5	32537	-	7.5000	7471A	PPB	12 May 2017 14:57:46	S	Std
10.0 - 1	510	44627	-	10.0000	7471A	PPB	12 May 2017 15:02:49	S	Std
ICV53 - 1	ICV53	16972	3.8301	-	7471A	PPB	12 May 2017 15:05:31	U	SMPL
ICB53 - 1	ICB53	-104	-0.0122	-	7471A	PPB	12 May 2017 15:07:46	U	SMPL
CCV93 - 1	CCV93	22747	5.1295	-	7471A	PPB	12 May 2017 15:10:02	U	SMPL
CCB93 - 1	CCB93	-118	-0.0154	-	7471A	PPB	12 May 2017 15:12:18	U	SMPL
CRA - 1	CRA	856	0.2038	-	7471A	PPB	12 May 2017 15:14:34	U	SMPL
HighStd - 1	HighStd	44630	10.0534	-	7471A	PPB	12 May 2017 15:16:49	U	SMPL
ChkStd - 1	ChkStd	31153	7.0210	-	7471A	PPB	12 May 2017 15:19:06	U	SMPL
PB98892BL - 1	PBS	-95	-0.0102	-	7471A	PPB	12 May 2017 15:21:23	U	SMPL
PB98892BS - 1	LCSS	15315	3.4572	-	7471A	PPB	12 May 2017 15:23:40	U	SMPL
I3059-01 - 1	JC-SWITCH-DRUM-SPILL(COMP)	35224	7.9370	-	7471A	PPB	12 May 2017 15:25:57	U	SMPL
I3065-01 - 1	PE-1	61390	13.8246	-	7471A	PPB	12 May 2017 15:28:14	U	SMPL
I3084-01 - 1	FK-PS-01-001	2143	0.4934	-	7471A	PPB	12 May 2017 15:30:31	U	SMPL
I3087-01 - 1	FK-PS-01_001	836	0.1993	-	7471A	PPB	12 May 2017 15:32:49	U	SMPL
I3087-01DUP - 1	FK-PS-01_001DUP	906	0.2151	-	7471A	PPB	12 May 2017 15:35:05	U	SMPL
CCV94 - 1	CCV94	22924	5.1693	-	7471A	PPB	12 May 2017 15:37:21	U	SMPL
CCB94 - 1	CCB94	-96	-0.0104	-	7471A	PPB	12 May 2017 15:39:37	U	SMPL
I3087-01MS - 1	FK-PS-01_001MS	16394	3.7000	-	7471A	PPB	12 May 2017 15:41:53	U	SMPL
I3087-01MSD - 1	FK-PS-01_001MSD	16160	3.6474	-	7471A	PPB	12 May 2017 15:44:09	U	SMPL
I3087-01LX5 - 1		132	0.0409	-	7471A	PPB	12 May 2017 15:49:19	U	SMPL
I3087-01A - 1		17370	3.9196	-	7471A	PPB	12 May 2017 15:51:35	U	SMPL
I3065-01DLX5 - 1		12290	2.7766	-	7471A	PPB	12 May 2017 15:57:12	U	SMPL
CCV95 - 1	CCV95	22677	5.1138	-	7471A	PPB	12 May 2017 15:59:28	U	SMPL
CCB95 - 1	CCB95	-104	-0.0122	-	7471A	PPB	12 May 2017 16:01:46	U	SMPL