



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87430

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/12/2017	16:00
2	CAL2	1	Water			7.00	05/12/2017	16:01
3	CAL3	1	Water			10.02	05/12/2017	16:02
4	ICV	1	Water			6.99	05/12/2017	16:05
5	CCV1	1	Water			2.01	05/12/2017	16:06
6	I3141-01	1	Solid	20.02	20	5.94	05/12/2017	16:07
7	I3141-01DUP	1	Solid	20.04	20	5.95	05/12/2017	16:09
8	I3144-01	1	Solid	20.01	20	5.78	05/12/2017	16:11
9	CCV2	1	Water			12.01	05/12/2017	16:12

LD 87436

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3141

WorkList ID : 98412

Date : 5/12/2017 3:48:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3141-01	pH	Cool 4 deg C	TULL02	H22	FK-PS-01_002	05/12/2017	9045C
	Solid	I3144-01	pH	Cool 4 deg C	TULL02	H51	FK-PS-01-002	05/12/2017	9045C

Date/Time 05-12-17 3:18 PM
 Received by: JD
 Relinquished by: JD

Date/Time 05-12-17 6:00 PM
 Received by: JD
 Relinquished by: JD