



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87435

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/12/2017	18:10
2	CAL2	1	Water			7.01	05/12/2017	18:11
3	CAL3	1	Water			10.04	05/12/2017	18:12
4	ICV	1	Water			7.01	05/12/2017	18:14
5	CCV1	1	Water			2.01	05/12/2017	18:15
6	I3147-07	1	Solid	20.03	20	6.77	05/12/2017	18:16
7	I3147-07DUP	1	Solid	20.04	20	6.78	05/12/2017	18:19
8	I3147-08	1	Solid	20.02	20	7.15	05/12/2017	18:20
9	I3147-09	1	Solid	20.05	20	7.58	05/12/2017	18:21
10	CCV2	1	Water			12.02	05/12/2017	18:22

LB 87435

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3147 WorkList ID : 98414 Date : 5/12/2017 5:57:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3147-07	Corrosivity	Cool 4 deg C	DEWB01	H33	TP-1COMP	05/11/2017	9045C
	Solid	I3147-08	Corrosivity	Cool 4 deg C	DEWB01	H33	TP-2COMP	05/12/2017	9045C
	Solid	I3147-09	Corrosivity	Cool 4 deg C	DEWB01	H33	TP-3COMP	05/12/2017	9045C

Date/Time 05-12-17 5:57:23 PM
Received by: JY
Relinquished by:

Date/Time 05-12-17 5:57:23 PM
Received by: CP
Relinquished by: JY