

7470A


LB 87448

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37093



A= 0.0000e+000

B= 2.6777e-004

C= 5.1030e-002

Rho= 0.9999387

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.040	0.040	-41	0.000	-41				
0.2	0.200	0.206	0.006	580	0.0 %	580				
2.5	2.500	2.438	-0.062	8915	0.0 %	8915				
5.0	5.000	4.972	-0.028	18379	0.0 %	18379				
7.5	7.500	7.559	0.059	28040	0.0 %	28040				
10.0	10.000	9.984	-0.016	37093	0.0 %	37093				

LB87448 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 May 2017 12:09:14

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	7470A	PPB	15 May 2017 12:13:45	S	Std
0.2 - 1	50.2	580	-	0.2000	7470A	PPB	15 May 2017 12:15:54	S	Std
2.5 - 1	52.5	8915	-	2.5000	7470A	PPB	15 May 2017 12:18:02	S	Std
5.0 - 1	55	18379	-	5.0000	7470A	PPB	15 May 2017 12:20:10	S	Std
7.5 - 1	57.5	28040	-	7.5000	7470A	PPB	15 May 2017 12:22:18	S	Std
10.0 - 1	516	37093	-	10.0000	7470A	PPB	15 May 2017 12:24:25	S	Std
ICV55 - 1	ICV55	14256	3.8684	-	7470A	PPB	15 May 2017 12:26:56	U	SMPL
ICB55 - 1	ICB55	-123	0.0181	-	7470A	PPB	15 May 2017 12:29:02	U	SMPL
CCV01 - 1	CCV01	18207	4.9264	-	7470A	PPB	15 May 2017 12:31:09	U	SMPL
CCB01 - 1	CCB01	-116	0.0200	-	7470A	PPB	15 May 2017 12:33:15	U	SMPL
CRA - 1	CRA	557	0.2002	-	7470A	PPB	15 May 2017 12:35:22	U	SMPL
HighStd - 1	HighStd	36806	9.9067	-	7470A	PPB	15 May 2017 12:37:28	U	SMPL
ChkStd - 1	ChkStd	25689	6.9299	-	7470A	PPB	15 May 2017 12:39:35	U	SMPL
PB98931BL - 1	PBW	-149	0.0111	-	7470A	PPB	15 May 2017 12:41:43	U	SMPL
PB98931BS - 1	LCSW	13532	3.6745	-	7470A	PPB	15 May 2017 12:43:51	U	SMPL
PB98931TB - 1	PB98882TB	-117	0.0197	-	7470A	PPB	15 May 2017 12:45:58	U	SMPL
I3092-07 - 1	SB-04-COMP	-55	0.0363	-	7470A	PPB	15 May 2017 12:48:06	U	SMPL
I3092-08 - 1	SB-05-COMP	-35	0.0417	-	7470A	PPB	15 May 2017 12:50:15	U	SMPL
I3092-09 - 1	SB-07-COMP	-44	0.0392	-	7470A	PPB	15 May 2017 12:52:23	U	SMPL
I3099-01 - 1	OVS-1	716	0.2428	-	7470A	PPB	15 May 2017 12:54:30	U	SMPL
CCV02 - 1	CCV02	18021	4.8766	-	7470A	PPB	15 May 2017 12:56:37	U	SMPL
CCB02 - 1	CCB02	-97	0.0251	-	7470A	PPB	15 May 2017 12:58:44	U	SMPL
I3108-01 - 1	IDW-GAC3-170510	-41	0.0401	-	7470A	PPB	15 May 2017 13:00:51	U	SMPL
I3108-01DUP - 1	IDW-GAC3-170510DUP	-29	0.0433	-	7470A	PPB	15 May 2017 13:02:58	U	SMPL
I3108-01MS - 1	IDW-GAC3-170510MS	13317	3.6170	-	7470A	PPB	15 May 2017 13:05:05	U	SMPL
I3108-01MSD - 1	IDW-GAC3-170510MSD	13310	3.6151	-	7470A	PPB	15 May 2017 13:07:12	U	SMPL
I3120-02 - 1	WP.RD-CORNELL-COMP	-70	0.0323	-	7470A	PPB	15 May 2017 13:09:19	U	SMPL
I3108-01LX5 - 1		-83	0.0288	-	7470A	PPB	15 May 2017 13:11:27	U	SMPL
I3108-01A - 1		13955	3.7878	-	7470A	PPB	15 May 2017 13:13:35	U	SMPL
CCV03 - 1	CCV03	17975	4.8643	-	7470A	PPB	15 May 2017 13:15:43	U	SMPL
CCB03 - 1	CCB03	-104	0.0232	-	7470A	PPB	15 May 2017 13:17:52	U	SMPL