



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 05/16/2017DATE & TIME IN: 05/15/2017 16:35OVEN TEMP: 98 °CRun Number: LB87457DATE & TIME OUT: 05/16/2017 12:45OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB87457BL	LB87457BL	90.5853	90.5853	100	90.5853	90.5853	90.5853	0	0
2	I3150-01	WS-1	90.8913	90.8913	100	90.9946	90.9946	90.9946	0.1033	1033
3	I3150-01DUP	WS-1DUP	94.1173	94.1173	100	94.2209	94.2209	94.2209	0.1036	1036
4	I3150-02	WS-2	95.234	95.234	100	95.6203	95.6203	95.6203	0.3863	3863

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LD8F457

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I3150

WorkList ID : 98457

Date : 5/15/2017 2:01:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3150-01-1	TS	Cool 4 deg C	TULL02	H61	WS-1	05/12/2017	SM2540 B
	Water	I3150-02-1	TS	Cool 4 deg C	TULL02	H61	WS-2	05/12/2017	SM2540 B

Date/Time 05-15-17 2:01 PM
 Received by: 711
 Relinquished by: CP

Date/Time 05-15-17 5:00 PM
 Received by: CP
 Relinquished by: JD