

7471A


LB 87458

INSTRUMENT ID: CVA

Linear

 μ Abs.:

42692



A= 0.0000e+000

B= 2.3102e-004

C= 1.1648e-002

Rho= 0.9994326

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-8	0.000	-8				
0.2	0.200	0.193	-0.007	787	0.0 %	787				
2.5	2.500	2.519	0.019	10854	0.0 %	10854				
5.0	5.000	4.863	-0.137	20999	0.0 %	20999				
7.5	7.500	7.740	0.240	33453	0.0 %	33453				
10.0	10.000	9.875	-0.125	42692	0.0 %	42692				

LB87458 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 May 2017 15:25:20

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.00007471A	PPB		15 May 2017 15:25:30	S	Std
0.2 - 1	50.2	787	-	0.20007471A	PPB		15 May 2017 15:27:48	S	Std
2.5 - 1	52.5	10854	-	2.50007471A	PPB		15 May 2017 15:30:05	S	Std
5.0 - 1	55	20999	-	5.00007471A	PPB		15 May 2017 15:32:23	S	Std
7.5 - 1	57.5	33453	-	7.50007471A	PPB		15 May 2017 15:34:40	S	Std
10.0 - 1	59	42692	-	10.00007471A	PPB		15 May 2017 15:39:33	S	Std
ICV56 - 1	ICV56	17651	4.0895	-7471A	PPB		15 May 2017 15:42:20	U	SMPL
ICB56 - 1	ICB56	-93	-0.0098	-7471A	PPB		15 May 2017 15:44:35	U	SMPL
CCV04 - 1	CCV04	20950	4.8516	-7471A	PPB		15 May 2017 15:46:51	U	SMPL
CCB04 - 1	CCB04	-113	-0.0145	-7471A	PPB		15 May 2017 15:49:07	U	SMPL
CRA - 1	CRA	752	0.1854	-7471A	PPB		15 May 2017 15:51:23	U	SMPL
HighStd - 1	HighStd	43024	9.9512	-7471A	PPB		15 May 2017 15:53:39	U	SMPL
ChkStd - 1	ChkStd	30982	7.1692	-7471A	PPB		15 May 2017 15:55:55	U	SMPL
PB98934BL - 1	PBS	-138	-0.0202	-7471A	PPB		15 May 2017 15:58:12	U	SMPL
PB98934BS - 1	LCSS	15444	3.5796	-7471A	PPB		15 May 2017 16:00:28	U	SMPL
I3037-03 - 1	HY-SB432A	6791	1.5805	-7471A	PPB		15 May 2017 16:02:45	U	SMPL
I3120-01 - 1	WPRD-CORNELL-COMP	10793	2.5051	-7471A	PPB		15 May 2017 16:05:02	U	SMPL
I3121-01 - 1	RO-1540-COMP	2730	0.6423	-7471A	PPB		15 May 2017 16:07:20	U	SMPL
I3122-01 - 1	RO-72-11997-72-12002-72-11979-COMP	535	0.1352	-7471A	PPB		15 May 2017 16:09:38	U	SMPL
I3122-01DUP - 1	RO-72-11997-72-12002-72-11979-COMP	513	0.1302	-7471A	PPB		15 May 2017 16:14:31	U	SMPL
CCV05 - 1	CCV05	20578	4.7657	-7471A	PPB		15 May 2017 16:16:46	U	SMPL
CCB05 - 1	CCB05	-65	-0.0034	-7471A	PPB		15 May 2017 16:19:02	U	SMPL
I3122-01MS - 1	RO-72-11997-72-12002-72-11979-COMPMS	16729	3.8765	-7471A	PPB		15 May 2017 16:21:18	U	SMPL
I3122-01MSD - 1	RO-72-11997-72-12002-72-11979-COMPMSD	16036	3.7164	-7471A	PPB		15 May 2017 16:23:34	U	SMPL
I3124-01 - 1	C91516-COMP	-68	-0.0041	-7471A	PPB		15 May 2017 16:25:50	U	SMPL
I3122-01LX5 - 1		126	0.0408	-7471A	PPB		15 May 2017 16:28:06	U	SMPL
I3122-01A - 1		17986	4.1668	-7471A	PPB		15 May 2017 16:30:23	U	SMPL
CCV06 - 1	CCV06	20701	4.7941	-7471A	PPB		15 May 2017 16:32:40	U	SMPL
CCB06 - 1	CCB06	-63	-0.0029	-7471A	PPB		15 May 2017 16:34:57	U	SMPL