



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87467

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/15/2017	17:00
2	CAL2	1	Water			7.00	05/15/2017	17:01
3	CAL3	1	Water			10.03	05/15/2017	17:02
4	ICV	1	Water			6.99	05/15/2017	17:03
5	CCV1	1	Water			2.01	05/15/2017	17:05
6	I3140-14	1	Solid	20.03	20	8.15	05/15/2017	17:06
7	I3140-14DUP	1	Solid	20.04	20	8.16	05/15/2017	17:07
8	I3140-18	1	Solid	20.02	20	8.16	05/15/2017	17:09
9	I3165-02	1	Solid	20.05	20	7.84	05/15/2017	17:10
10	I3169-01	1	Solid	20.04	20	6.95	05/15/2017	17:11
11	CCV2	1	Water			12.01	05/15/2017	17:12

15787467

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I3165

WorkList ID : 98470

Date : 5/15/2017 4:31:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/24/2017	Solid	I3140-14	Corrosivity	Cool 4 deg C	DAYE02	H43	TP-17-7	05/10/2017	9045C
05/24/2017	Solid	I3140-18	Corrosivity	Cool 4 deg C	DAYE02	H43	TP-25-6	05/10/2017	9045C
	Solid	I3165-02	Corrosivity	Cool 4 deg C	AREC01	H23	WC-1	05/13/2017	9045C
	Solid	I3169-01	Corrosivity	Cool 4 deg C	CLEA11	H21	OWS-3	05/12/2017	9045C

Date/Time 05-15-17 4:31 PM
 Received by: JD
 Relinquished by: RC

Date/Time 05-15-17 7:00 AM
 Received by: R-C
 Relinquished by: JD