

7470A


LB 87476

INSTRUMENT ID: CY2

Linear

 μ Abs.:

35362



A= 0.0000e+000

B= 2.8330e-004

C= 2.7386e-002

Rho= 0.9999480

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	-85	0.000	-85				
0.2	0.200	0.202	0.002	615	0.0 %	615				
2.5	2.500	2.544	0.044	8882	0.0 %	8882				
5.0	5.000	4.947	-0.053	17365	0.0 %	17365				
7.5	7.500	7.459	-0.041	26232	0.0 %	26232				
10.0	10.000	10.046	0.046	35362	0.0 %	35362				

LB87476 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 May 2017 10:01:43

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-85	-	0.0000	7470A	PPB	16 May 2017 10:09:08	S	Std
0.2 - 1	50.2	615	-	0.2000	7470A	PPB	16 May 2017 10:11:16	S	Std
2.5 - 1	52.5	8882	-	2.5000	7470A	PPB	16 May 2017 10:13:25	S	Std
5.0 - 1	35	17365	-	5.0000	7470A	PPB	16 May 2017 10:18:54	S	Std
7.5 - 1	57.5	26232	-	7.5000	7470A	PPB	16 May 2017 10:21:01	S	Std
10.0 - 1	510	35362	-	10.0000	7470A	PPB	16 May 2017 10:25:36	S	Std
ICV57 - 1	ICV57	13964	3.9834	-	7470A	PPB	16 May 2017 10:28:06	U	SMPL
ICB57 - 1	ICB57	-134	-0.0106	-	7470A	PPB	16 May 2017 10:30:12	U	SMPL
CCV07 - 1	CCV07	17638	5.0243	-	7470A	PPB	16 May 2017 10:32:18	U	SMPL
CCB07 - 1	CCB07	-128	-0.0089	-	7470A	PPB	16 May 2017 10:34:25	U	SMPL
CRA - 1	CRA	641	0.2090	-	7470A	PPB	16 May 2017 10:36:31	U	SMPL
HighStd - 1	HighStd	35269	10.0192	-	7470A	PPB	16 May 2017 10:38:38	U	SMPL
ChkStd - 1	ChkStd	25967	7.3839	-	7470A	PPB	16 May 2017 10:40:45	U	SMPL
PB98971BL - 1	PBW	-173	-0.0216	-	7470A	PPB	16 May 2017 10:42:53	U	SMPL
PB98971BS - 1	LCSW	14714	4.1959	-	7470A	PPB	16 May 2017 10:45:01	U	SMPL
PB98971TB - 1	PB98926TB	-127	-0.0086	-	7470A	PPB	16 May 2017 10:47:08	U	SMPL
I3121-03 - 1	RO-1540-COMP	-65	0.0090	-	7470A	PPB	16 May 2017 10:49:17	U	SMPL
I3122-02 - 1	RO-72-11997-72-12002-72-11979-COMP	-49	0.0135	-	7470A	PPB	16 May 2017 10:51:25	U	SMPL
I3123-01 - 1	DEBRIS-DRUMS-COMP	-22	0.0212	-	7470A	PPB	16 May 2017 10:53:33	U	SMPL
I3147-07 - 1	TP-1COMP	-42	0.0155	-	7470A	PPB	16 May 2017 10:55:40	U	SMPL
CCV08 - 1	CCV08	18033	5.1362	-	7470A	PPB	16 May 2017 11:00:23	U	SMPL
CCB08 - 1	CCB08	-142	-0.0128	-	7470A	PPB	16 May 2017 11:02:30	U	SMPL
I3147-08 - 1	TP-2COMP	-54	0.0121	-	7470A	PPB	16 May 2017 11:04:36	U	SMPL
I3147-09 - 1	TP-3COMP	-59	0.0107	-	7470A	PPB	16 May 2017 11:06:42	U	SMPL
I3147-09DUP - 1	TP-3COMP DUP	-51	0.0129	-	7470A	PPB	16 May 2017 11:08:49	U	SMPL
I3147-09MS - 1	TP-3COMPMS	13224	3.7738	-	7470A	PPB	16 May 2017 11:10:56	U	SMPL
I3147-09MSD - 1	TP-3COMPMSD	13547	3.8653	-	7470A	PPB	16 May 2017 11:13:04	U	SMPL
PB98981BL - 1	PBW	-54	0.0121	-	7470A	PPB	16 May 2017 11:15:12	U	SMPL
PB98981BS - 1	LCSW	13337	3.8058	-	7470A	PPB	16 May 2017 11:17:20	U	SMPL
I3140-01 - 1	TANK-1	22482	6.3966	-	7470A	PPB	16 May 2017 11:19:28	U	SMPL
I3140-02 - 1	MW-C	1615	0.4849	-	7470A	PPB	16 May 2017 11:21:37	U	SMPL
I3140-03 - 1	MW-A	18380	5.2345	-	7470A	PPB	16 May 2017 11:23:44	U	SMPL
CCV09 - 1	CCV09	17401	4.9571	-	7470A	PPB	16 May 2017 11:25:51	U	SMPL
CCB09 - 1	CCB09	-89	0.0022	-	7470A	PPB	16 May 2017 11:27:58	U	SMPL
I3152-01 - 1	MW-6A-R1	-66	0.0087	-	7470A	PPB	16 May 2017 11:30:05	U	SMPL
I3152-01DUP - 1	MW-6A-R1DUP	-85	0.0033	-	7470A	PPB	16 May 2017 11:32:11	U	SMPL
I3152-01MS - 1	MW-6A-R1MS	14166	4.0406	-	7470A	PPB	16 May 2017 11:34:17	U	SMPL
I3152-01MSD - 1	MW-6A-R1MSD	14085	4.0177	-	7470A	PPB	16 May 2017 11:36:24	U	SMPL
I3152-02 - 1	MW-6B-R2	-112	-0.0043	-	7470A	PPB	16 May 2017 11:38:31	U	SMPL
I3152-03 - 1	MW-4A-R1	35	0.0373	-	7470A	PPB	16 May 2017 11:40:39	U	SMPL
I3152-04 - 1	MW-5A-R1	-89	0.0022	-	7470A	PPB	16 May 2017 11:42:46	U	SMPL
I3152-05 - 1	MW-5B-R2	-84	0.0036	-	7470A	PPB	16 May 2017 11:44:54	U	SMPL
I3152-06 - 1	FB-R1-051017	-85	0.0033	-	7470A	PPB	16 May 2017 11:47:03	U	SMPL
I3152-07 - 1	MW-3D-R1	-42	0.0155	-	7470A	PPB	16 May 2017 11:49:10	U	SMPL
CCV10 - 1	CCV10	17614	5.0175	-	7470A	PPB	16 May 2017 11:51:17	U	SMPL
CCB10 - 1	CCB10	-96	0.0002	-	7470A	PPB	16 May 2017 11:53:24	U	SMPL
I3152-08 - 1	MW-3-R1	-83	0.0039	-	7470A	PPB	16 May 2017 11:55:31	U	SMPL
I3152-09 - 1	MW-2D-R1	-67	0.0084	-	7470A	PPB	16 May 2017 11:57:38	U	SMPL
I3152-10 - 1	MW-2-R1	-4	0.0263	-	7470A	PPB	16 May 2017 11:59:45	U	SMPL
I3152-11 - 1	FB-R1-051117	-83	0.0039	-	7470A	PPB	16 May 2017 12:01:52	U	SMPL
I3152-12 - 1	MW-1-R1	-73	0.0067	-	7470A	PPB	16 May 2017 12:04:00	U	SMPL
I3152-13 - 1	MW-1D-R1	-60	0.0104	-	7470A	PPB	16 May 2017 12:06:07	U	SMPL
I3152-14 - 1	MW-4D-R1	-65	0.0090	-	7470A	PPB	16 May 2017 12:08:15	U	SMPL
I3152-15 - 1	MW-5D-R1	-84	0.0036	-	7470A	PPB	16 May 2017 12:10:23	U	SMPL
I3152-16 - 1	FB-R1-051217	-80	0.0047	-	7470A	PPB	16 May 2017 12:12:31	U	SMPL
I3152-17 - 1	DUP-R1	-78	0.0053	-	7470A	PPB	16 May 2017 12:14:39	U	SMPL
CCV11 - 1	CCV11	17475	4.9781	-	7470A	PPB	16 May 2017 12:16:46	U	SMPL
CCB11 - 1	CCB11	-103	-0.0018	-	7470A	PPB	16 May 2017 12:18:53	U	SMPL
I3147-09LX5 - 1		-39	0.0163	-	7470A	PPB	16 May 2017 12:21:00	U	SMPL
I3147-09A - 1		14734	4.2016	-	7470A	PPB	16 May 2017 12:23:07	U	SMPL
I3152-01Lx5 - 1		-97	-0.0001	-	7470A	PPB	16 May 2017 12:25:15	U	SMPL
I3152-01A - 1		15917	4.5367	-	7470A	PPB	16 May 2017 12:27:22	U	SMPL
CCV12 - 1	CCV12	17915	5.1027	-	7470A	PPB	16 May 2017 12:29:29	U	SMPL
CCB12 - 1	CCB12	-120	-0.0066	-	7470A	PPB	16 May 2017 12:31:37	U	SMPL