

7471A


LB87481

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41237



A= 0.0000e+000

B= 2.4185e-004

C= 1.7572e-002

Rho= 0.9999686

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	-28	0.000	-28				
0.2	0.200	0.209	0.009	790	0.0 %	790				
2.5	2.500	2.493	-0.007	10234	0.0 %	10234				
5.0	5.000	4.950	-0.050	20394	0.0 %	20394				
7.5	7.500	7.548	0.048	31136	0.0 %	31136				
10.0	10.000	9.991	-0.009	41237	0.0 %	41237				

LB87481 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 16 May 2017 12:42:48

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.0000	7471A	PPB	16 May 2017 12:43:59	S	Std
0.2 - 1	50.2	790	-	0.2000	7471A	PPB	16 May 2017 12:46:16	S	Std
2.5 - 1	52.5	10234	-	2.5000	7471A	PPB	16 May 2017 12:48:34	S	Std
5.0 - 1	55	20394	-	5.0000	7471A	PPB	16 May 2017 12:50:52	S	Std
7.5 - 1	57.5	31136	-	7.5000	7471A	PPB	16 May 2017 12:55:44	S	Std
10.0 - 1	59	41237	-	10.0000	7471A	PPB	16 May 2017 12:58:00	S	Std
ICV58 - 1	ICV58	15901	3.8632	-7471A	PPB	16 May 2017 13:00:45	U	SMPL	
ICB58 - 1	ICB58	-88	-0.0037	-7471A	PPB	16 May 2017 13:03:00	U	SMPL	
CCV13 - 1	CCV13	19693	4.7802	-7471A	PPB	16 May 2017 13:05:16	U	SMPL	
CCB13 - 1	CCB13	-83	-0.0025	-7471A	PPB	16 May 2017 13:07:32	U	SMPL	
CRA - 1	CRA	759	0.2011	-7471A	PPB	16 May 2017 13:09:48	U	SMPL	
HighStd - 1	HighStd	41758	10.1165	-7471A	PPB	16 May 2017 13:12:04	U	SMPL	
ChkStd - 1	ChkStd	30405	7.3709	-7471A	PPB	16 May 2017 13:14:20	U	SMPL	
PB98982BL - 1	PBS	-138	-0.0158	-7471A	PPB	16 May 2017 13:16:37	U	SMPL	
PB98982BS - 1	LCSS	14520	3.5292	-7471A	PPB	16 May 2017 13:18:54	U	SMPL	
I3139-01 - 1	E2-U10-SSW	13549	3.2943	-7471A	PPB	16 May 2017 13:21:10	U	SMPL	
I3139-02 - 1	E2-U10-ESW	21455	5.2064	-7471A	PPB	16 May 2017 13:23:28	U	SMPL	
I3140-06 - 1	BERM-1-2	12520	3.0455	-7471A	PPB	16 May 2017 13:25:45	U	SMPL	
I3140-07 - 1	BERM-2-2	21493	5.2155	-7471A	PPB	16 May 2017 13:28:03	U	SMPL	
I3140-09 - 1	TB-12-6	27222	6.6011	-7471A	PPB	16 May 2017 13:30:19	U	SMPL	
CCV14 - 1	CCV14	19890	4.8279	-7471A	PPB	16 May 2017 13:32:35	U	SMPL	
CCB14 - 1	CCB14	-80	-0.0018	-7471A	PPB	16 May 2017 13:34:51	U	SMPL	
I3140-10 - 1	TP-3-3	6184	1.5131	-7471A	PPB	16 May 2017 13:37:08	U	SMPL	
I3140-11 - 1	TP-4-3	24931	6.0470	-7471A	PPB	16 May 2017 13:39:24	U	SMPL	
I3140-14 - 1	TP-17-7	22543	5.4695	-7471A	PPB	16 May 2017 13:41:40	U	SMPL	
I3140-15 - 1	TP-25-6	14496	3.5234	-7471A	PPB	16 May 2017 13:43:56	U	SMPL	
I3140-16 - 1	TP-30-3	69611	16.8526	-7471A	PPB	16 May 2017 13:46:13	U	SMPL	
I3141-01 - 1	FK-PS-01_002	693	0.1852	-7471A	PPB	16 May 2017 13:48:30	U	SMPL	
I3144-01 - 1	FK-PS-01-002	1656	0.4181	-7471A	PPB	16 May 2017 13:50:47	U	SMPL	
I3147-02 - 1	TP-1COMP	374	0.1080	-7471A	PPB	16 May 2017 13:53:05	U	SMPL	
I3147-02DUP - 1	TP-1COMPDPUP	351	0.1025	-7471A	PPB	16 May 2017 13:55:23	U	SMPL	
I3147-02MS - 1	TP-1COMPMS	15354	3.7309	-7471A	PPB	16 May 2017 13:57:39	U	SMPL	
CCV15 - 1	CCV15	20003	4.8552	-7471A	PPB	16 May 2017 13:59:55	U	SMPL	
CCB15 - 1	CCB15	-60	0.0031	-7471A	PPB	16 May 2017 14:02:12	U	SMPL	
I3147-02MSD - 1	TP-1COMPMSD	15523	3.7717	-7471A	PPB	16 May 2017 14:04:28	U	SMPL	
I3147-04 - 1	TP-2COMP	290	0.0877	-7471A	PPB	16 May 2017 14:06:45	U	SMPL	
I3147-06 - 1	TP-3COMP	660	0.1772	-7471A	PPB	16 May 2017 14:09:01	U	SMPL	
I3151-01 - 1	NM10-COMP-0-6	4547	1.1172	-7471A	PPB	16 May 2017 14:11:18	U	SMPL	
I3151-02 - 1	NM10-COMP-6-18	16775	4.0745	-7471A	PPB	16 May 2017 14:13:34	U	SMPL	
PB98983BL - 1	PBS	-93	-0.0049	-7471A	PPB	16 May 2017 14:15:51	U	SMPL	
PB98983BS - 1	LCSS	14266	3.4677	-7471A	PPB	16 May 2017 14:18:09	U	SMPL	
I3157-01 - 1	NORTH-B	1340	0.3416	-7471A	PPB	16 May 2017 14:20:26	U	SMPL	
I3157-02 - 1	SOUTH-B	5931	1.4520	-7471A	PPB	16 May 2017 14:22:44	U	SMPL	
I3157-03 - 1	FLOOR-B	3478	0.8587	-7471A	PPB	16 May 2017 14:25:01	U	SMPL	
CCV16 - 1	CCV16	20010	4.8569	-7471A	PPB	16 May 2017 14:27:17	U	SMPL	
CCB16 - 1	CCB16	-73	-0.0001	-7471A	PPB	16 May 2017 14:29:33	U	SMPL	
I3157-04 - 1	NORTH-C	1591	0.4023	-7471A	PPB	16 May 2017 14:31:50	U	SMPL	
I3157-05 - 1	SOUTH-C	1146	0.2947	-7471A	PPB	16 May 2017 14:34:06	U	SMPL	
I3157-06 - 1	WEST	562	0.1535	-7471A	PPB	16 May 2017 14:36:22	U	SMPL	
I3157-06DUP - 1	WESTDUP	630	0.1699	-7471A	PPB	16 May 2017 14:41:32	U	SMPL	
I3157-06MS - 1	WESTMS	16003	3.8878	-7471A	PPB	16 May 2017 14:43:48	U	SMPL	
I3157-06MSD - 1	WESTMSD	16096	3.9103	-7471A	PPB	16 May 2017 14:46:05	U	SMPL	
I3160-01 - 1	S-WC-2	436	0.1230	-7471A	PPB	16 May 2017 14:48:22	U	SMPL	
I3160-02 - 1	S-WC-3	6841	1.6720	-7471A	PPB	16 May 2017 14:50:40	U	SMPL	
I3160-03 - 1	S-WC-4	10951	2.6660	-7471A	PPB	16 May 2017 14:52:57	U	SMPL	
I3165-01 - 1	WC-1	5590	1.3695	-7471A	PPB	16 May 2017 14:55:14	U	SMPL	
CCV17 - 1	CCV17	20636	5.0083	-7471A	PPB	16 May 2017 14:57:30	U	SMPL	
CCB17 - 1	CCB17	-82	-0.0023	-7471A	PPB	16 May 2017 14:59:47	U	SMPL	
I3167-01 - 1	MEADOW-SOIL	5112	1.2539	-7471A	PPB	16 May 2017 15:02:03	U	SMPL	
I3167-02 - 1	HORTICULTURAL-SUB-SOIL	3011	0.7458	-7471A	PPB	16 May 2017 15:04:19	U	SMPL	
I3170-01 - 1	POSTEX-32-20170515	10260	2.4989	-7471A	PPB	16 May 2017 15:06:36	U	SMPL	
I3170-02 - 1	POSTEX-33-20170515	19161	4.6516	-7471A	PPB	16 May 2017 15:08:52	U	SMPL	
I3147-02LX5 - 1		-4	0.0166	-7471A	PPB	16 May 2017 15:11:09	U	SMPL	
I3147-02A - 1		17201	4.1775	-7471A	PPB	16 May 2017 15:13:26	U	SMPL	
I3157-06LX5 - 1		29	0.0246	-7471A	PPB	16 May 2017 15:15:43	U	SMPL	
I3157-06A - 1		17598	4.2736	-7471A	PPB	16 May 2017 15:18:00	U	SMPL	
I3140-16DLX5 - 1		13044	3.1722	-7471A	PPB	16 May 2017 15:20:19	U	SMPL	
CCV18 - 1	CCV18	20636	5.0083	-7471A	PPB	16 May 2017 15:22:35	U	SMPL	
CCB18 - 1	CCB18	-67	0.0014	-7471A	PPB	16 May 2017 15:24:52	U	SMPL	