

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB87483

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54973
calibration std. phosphate 0.5 ppm	WP54972
calibration std. phosphate 0.3 ppm	WP54970
calibration std. 0 ppm	WP54963
phosphate ICV std.	WP54975
calibration std. phosphate 0.1 ppm	WP54968
calibration std. phosphate 0.05 ppm	WP54966
Combined reagent	WP54977
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate CCV std.	WP54974
phosphate ICV std.	WP54975

Intercept: -0.0008Slope: 0.6491Regression: 0.999804

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/16/2017	13:34
2	CAL2	0.05	1	50	50	0.032	05/16/2017	13:34
3	CAL3	0.10	1	50	50	0.065	05/16/2017	13:35
4	CAL4	0.30	1	50	50	0.186	05/16/2017	13:35
5	CAL5	0.50	1	50	50	0.331	05/16/2017	13:36
6	CAL6	1.00	1	50	50	0.647	05/16/2017	13:36



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB87483

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.315	0.487	05/16/2017	13:37
2	ICB		1	50	50	0.000	0.001	05/16/2017	13:37
3	CCV1	0.50	1	50	50	0.326	0.503	05/16/2017	13:38
4	CCB1		1	50	50	0.000	0.001	05/16/2017	13:38
5	PB99006BL		1	50	50	0.000	0.001	05/16/2017	13:38
6	PB99006BS	0.50	1	50	50	0.313	0.483	05/16/2017	13:39
7	I3158-01		1	50	50	0.068	0.106	05/16/2017	13:40
8	I3158-01DUP		1	50	50	0.070	0.109	05/16/2017	13:40
9	I3158-01MS	0.50	1	50	50	0.371	0.573	05/16/2017	13:41
10	I3158-01MSD	0.50	1	50	50	0.372	0.574	05/16/2017	13:41
11	CCV2	0.50	1	50	50	0.327	0.505	05/16/2017	13:42
12	CCB2		1	50	50	0.000	0.001	05/16/2017	13:42