



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 05/17/2017DATE & TIME IN: 05/16/2017 16:40OVEN TEMP: 98 °CRun Number: LB87489DATE & TIME OUT: 05/17/2017 12:45OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB87489BL	LB87489BL	78.4019	78.4019	100	78.4019	78.4019	78.4019	0	0
2	I3171-01	WS-4	83.5082	83.5082	100	83.8566	83.8566	83.8566	0.3484	3484
3	I3171-01DUP	WS-4DUP	73.6684	73.6684	100	74.0172	74.0172	74.0172	0.3488	3488
4	I3171-02	WS-5	79.5084	79.5084	100	80.969	80.969	80.969	1.4606	14606

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB 87489

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I3171

WorkList ID : 98515

Date : 5/16/2017 2:39:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3171-01-F	TS	Cool 4 deg C	TULL02	H32	WS-4	05/15/2017	SM2540 B
	Water	I3171-02-F	TS	Cool 4 deg C	TULL02	H32	WS-5	05/15/2017	SM2540 B

Date/Time 05-16-17 2:30 PM
 Received by: JP
 Relinquished by: JP

Date/Time 05-16-17 5:00 PM
 Received by: CP
 Relinquished by: JP