

7470A


LB 87506

INSTRUMENT ID : CV1

Linear

 $\mu$  Abs.:

3490E



A= 0.0000e+000

B= 2.8427e-004

C= -3.8676e-002

Rho= 0.9997625

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | -0.018 | -0.018 | 74    | 0.000      | 74    |       |       |       |       |
| 0.2    | 0.200  | 0.180  | -0.020 | 768   | 0.0 %      | 768   |       |       |       |       |
| 2.5    | 2.500  | 2.513  | 0.013  | 8978  | 0.0 %      | 8978  |       |       |       |       |
| 5.0    | 5.000  | 4.985  | -0.015 | 17672 | 0.0 %      | 17672 |       |       |       |       |
| 7.5    | 7.500  | 7.656  | 0.156  | 27067 | 0.0 %      | 27067 |       |       |       |       |
| 10.0   | 10.000 | 9.884  | -0.116 | 34906 | 0.0 %      | 34906 |       |       |       |       |

## LB87506 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 May 2017 11:42:02

| Sample ID       | Extended ID      | $\mu$ Abs. | Conc.   | Std Conc | Method | Units                | Date                 | Type | Type |
|-----------------|------------------|------------|---------|----------|--------|----------------------|----------------------|------|------|
| 0.0 - 1         | SO               | 74         | -       | 0.0000   | 7470A  | PPB                  | 17 May 2017 11:52:47 | S    | Std  |
| 0.2 - 1         | SO.2             | 768        | -       | 0.2000   | 7470A  | PPB                  | 17 May 2017 11:54:55 | S    | Std  |
| 2.5 - 1         | SO.5             | 8978       | -       | 2.5000   | 7470A  | PPB                  | 17 May 2017 11:57:04 | S    | Std  |
| 5.0 - 1         | SO.5             | 17672      | -       | 5.0000   | 7470A  | PPB                  | 17 May 2017 11:59:12 | S    | Std  |
| 7.5 - 1         | SO.5             | 27067      | -       | 7.5000   | 7470A  | PPB                  | 17 May 2017 12:01:20 | S    | Std  |
| 10.0 - 1        | SO               | 34906      | -       | 10.0000  | 7470A  | PPB                  | 17 May 2017 12:03:28 | S    | Std  |
| ICV59 - 1       | ICV59            | 13886      | 3.9087  | -7470A   | PPB    | 17 May 2017 12:06:11 | U                    | SMPL |      |
| ICB59 - 1       | ICB59            | -13        | -0.0424 | -7470A   | PPB    | 17 May 2017 12:08:17 | U                    | SMPL |      |
| CCV19 - 1       | CCV19            | 17827      | 5.0290  | -7470A   | PPB    | 17 May 2017 12:10:23 | U                    | SMPL |      |
| CCB19 - 1       | CCB19            | 2          | -0.0381 | -7470A   | PPB    | 17 May 2017 12:12:30 | U                    | SMPL |      |
| CRA - 1         | CRA              | 759        | 0.1771  | -7470A   | PPB    | 17 May 2017 12:14:37 | U                    | SMPL |      |
| HighStd - 1     | HighStd          | 35370      | 10.0159 | -7470A   | PPB    | 17 May 2017 12:16:43 | U                    | SMPL |      |
| ChkStd - 1      | ChkStd           | 25779      | 7.2895  | -7470A   | PPB    | 17 May 2017 12:18:50 | U                    | SMPL |      |
| PB99008BL - 1   | PBW              | -3         | -0.0395 | -7470A   | PPB    | 17 May 2017 12:20:58 | U                    | SMPL |      |
| PB99008BS - 1   | LCSW             | 13500      | 3.7989  | -7470A   | PPB    | 17 May 2017 12:23:05 | U                    | SMPL |      |
| PB99008TB - 1   | PB98958TB        | 22         | -0.0324 | -7470A   | PPB    | 17 May 2017 12:25:13 | U                    | SMPL |      |
| I3140-18 - 1    | TP-25-6          | 75         | -0.0174 | -7470A   | PPB    | 17 May 2017 12:27:21 | U                    | SMPL |      |
| I3165-02 - 1    | WC-1             | 75         | -0.0174 | -7470A   | PPB    | 17 May 2017 12:29:29 | U                    | SMPL |      |
| I3165-02DUP - 1 | WC-1DUP          | 89         | -0.0134 | -7470A   | PPB    | 17 May 2017 12:31:38 | U                    | SMPL |      |
| I3165-02MS - 1  | WC-1MS           | 12893      | 3.6264  | -7470A   | PPB    | 17 May 2017 12:33:45 | U                    | SMPL |      |
| CCV20 - 1       | CCV20            | 17731      | 5.0017  | -7470A   | PPB    | 17 May 2017 12:35:52 | U                    | SMPL |      |
| CCB20 - 1       | CCB20            | -12        | -0.0421 | -7470A   | PPB    | 17 May 2017 12:37:58 | U                    | SMPL |      |
| I3165-02MSD - 1 | WC-1MSD          | 13580      | 3.8217  | -7470A   | PPB    | 17 May 2017 12:40:05 | U                    | SMPL |      |
| I3169-01 - 1    | OWS-3            | 1213       | 0.3061  | -7470A   | PPB    | 17 May 2017 12:42:12 | U                    | SMPL |      |
| PB99009BL - 1   | PBW              | 41         | -0.0270 | -7470A   | PPB    | 17 May 2017 12:44:18 | U                    | SMPL |      |
| PB99009BS - 1   | LCSW             | 13693      | 3.8538  | -7470A   | PPB    | 17 May 2017 12:46:25 | U                    | SMPL |      |
| I3159-01 - 1    | MW-3-20170515    | 16         | -0.0341 | -7470A   | PPB    | 17 May 2017 12:48:33 | U                    | SMPL |      |
| I3159-01DUP - 1 | MW-3-20170515DUP | 95         | -0.0117 | -7470A   | PPB    | 17 May 2017 12:50:40 | U                    | SMPL |      |
| I3159-01MS - 1  | MW-3-20170515MS  | 15277      | 4.3041  | -7470A   | PPB    | 17 May 2017 12:52:49 | U                    | SMPL |      |
| I3159-01MSD - 1 | MW-3-20170515MSD | 14556      | 4.0991  | -7470A   | PPB    | 17 May 2017 12:54:57 | U                    | SMPL |      |
| I3159-02 - 1    | MW-3-20170515    | -7         | -0.0407 | -7470A   | PPB    | 17 May 2017 12:57:05 | U                    | SMPL |      |
| CCV21 - 1       | CCV21            | 17585      | 4.9602  | -7470A   | PPB    | 17 May 2017 12:59:12 | U                    | SMPL |      |
| CCB21 - 1       | CCB21            | 8          | -0.0364 | -7470A   | PPB    | 17 May 2017 13:01:19 | U                    | SMPL |      |
| I3165-02LX5 - 1 |                  | 106        | -0.0085 | -7470A   | PPB    | 17 May 2017 13:03:26 | U                    | SMPL |      |
| I3165-02A - 1   |                  | 15210      | 4.2850  | -7470A   | PPB    | 17 May 2017 13:05:33 | U                    | SMPL |      |
| I3159-01LX5 - 1 |                  | 33         | -0.0293 | -7470A   | PPB    | 17 May 2017 13:07:40 | U                    | SMPL |      |
| I3159-01A - 1   |                  | 16238      | 4.5773  | -7470A   | PPB    | 17 May 2017 13:09:47 | U                    | SMPL |      |
| CCV22 - 1       | CCV22            | 18025      | 5.0853  | -7470A   | PPB    | 17 May 2017 13:11:54 | U                    | SMPL |      |
| CCB22 - 1       | CCB22            | -3         | -0.0395 | -7470A   | PPB    | 17 May 2017 13:14:01 | U                    | SMPL |      |