

245.1


LB87508

INSTRUMENT ID: CY1

Linear

 μ Abs.:

32607



A= 0.0000e+000

B= 3.0518e-004

C= -2.9078e-002

Rho= 0.9997609

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	68	0.000	68				
0.2	0.200	0.198	-0.002	744	0.0 %	744				
2.5	2.500	2.531	0.031	8390	0.0 %	8390				
5.0	5.000	4.906	-0.094	16170	0.0 %	16170				
7.5	7.500	7.652	0.152	25168	0.0 %	25168				
10.0	10.000	9.922	-0.078	32607	0.0 %	32607				

LB87508 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 May 2017 14:01:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	68	-	0.0000	245.1	PPB	17 May 2017 14:12:35	S	Std
0.2 - 1	50.2	744	-	0.2000	245.1	PPB	17 May 2017 14:14:44	S	Std
2.5 - 1	52.5	8390	-	2.5000	245.1	PPB	17 May 2017 14:16:52	S	Std
5.0 - 1	55	16170	-	5.0000	245.1	PPB	17 May 2017 14:19:00	S	Std
7.5 - 1	57.5	25168	-	7.5000	245.1	PPB	17 May 2017 14:21:08	S	Std
10.0 - 1	510	32607	-	10.0000	245.1	PPB	17 May 2017 14:25:55	S	Std
ICV60 - 1	ICV60	13221	4.0057	-245.1		PPB	17 May 2017 14:28:27	U	SMPL
ICB60 - 1	ICB60	36	-0.0181	-245.1		PPB	17 May 2017 14:30:33	U	SMPL
CCV23 - 1	CCV23	16209	4.9175	-245.1		PPB	17 May 2017 14:32:39	U	SMPL
CCB23 - 1	CCB23	38	-0.0175	-245.1		PPB	17 May 2017 14:34:46	U	SMPL
CRA - 1	CRA	760	0.2029	-245.1		PPB	17 May 2017 14:36:53	U	SMPL
HighStd - 1	HighStd	32301	9.8284	-245.1		PPB	17 May 2017 14:38:59	U	SMPL
ChkStd - 1	ChkStd	23920	7.2707	-245.1		PPB	17 May 2017 14:41:05	U	SMPL
PB99007BL - 1	PBW	12	-0.0254	-245.1		PPB	17 May 2017 14:43:13	U	SMPL
PB99007BS - 1	LCSW	14198	4.3038	-245.1		PPB	17 May 2017 14:45:20	U	SMPL
I3150-01 - 1	WS-1	722	0.1913	-245.1		PPB	17 May 2017 14:47:28	U	SMPL
I3150-02 - 1	WS-2	315	0.0671	-245.1		PPB	17 May 2017 14:56:16	U	SMPL
I3150-02DUP - 1	WS-2DUP	289	0.0591	-245.1		PPB	17 May 2017 14:58:23	U	SMPL
I3150-02MS - 1	WS-2MS	15785	4.7881	-245.1		PPB	17 May 2017 15:00:32	U	SMPL
I3150-02MSD - 1	WS-2MSD	15419	4.6764	-245.1		PPB	17 May 2017 15:02:39	U	SMPL
CCV24 - 1	CCV24	16093	4.8821	-245.1		PPB	17 May 2017 15:04:46	U	SMPL
CCB24 - 1	CCB24	45	-0.0153	-245.1		PPB	17 May 2017 15:06:53	U	SMPL
I3171-01 - 1	WS-4	138	0.0130	-245.1		PPB	17 May 2017 15:08:59	U	SMPL
I3171-02 - 1	WS-5	2252	0.6582	-245.1		PPB	17 May 2017 15:11:05	U	SMPL
I3150-02Lx5 - 1		105	0.0030	-245.1		PPB	17 May 2017 15:13:12	U	SMPL
I3150-02A - 1		17439	5.2929	-245.1		PPB	17 May 2017 15:15:19	U	SMPL
CCV25 - 1	CCV25	16337	4.9566	-245.1		PPB	17 May 2017 15:17:27	U	SMPL
CCB25 - 1	CCB25	34	-0.0187	-245.1		PPB	17 May 2017 15:19:35	U	SMPL