


7471A

L387513
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42076



A= 0.0000e+000

B= 2.3525e-004

C= -1.8666e-002

Rho= 0.9997384

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	24	0.000	24				
0.2	0.200	0.191	-0.009	893	0.0 %	893				
2.5	2.500	2.485	-0.015	10644	0.0 %	10644				
5.0	5.000	4.990	-0.010	21292	0.0 %	21292				
7.5	7.500	7.666	0.166	32668	0.0 %	32668				
10.0	10.000	9.880	-0.120	42076	0.0 %	42076				

LB87513 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 17 May 2017 16:00:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	24	-	0.0000	7471A	PPB	17 May 2017 16:02:24	S	Std
0.2 - 1	S0.2	893	-	0.2000	7471A	PPB	17 May 2017 16:04:41	S	Std
2.5 - 1	S2.5	10644	-	2.5000	7471A	PPB	17 May 2017 16:06:59	S	Std
5.0 - 1	S5	21292	-	5.0000	7471A	PPB	17 May 2017 16:09:17	S	Std
7.5 - 1	S7.5	32668	-	7.5000	7471A	PPB	17 May 2017 16:11:34	S	Std
10.0 - 1	S10	42076	-	10.0000	7471A	PPB	17 May 2017 16:13:50	S	Std
ICV61 - 1	ICV61	16114	3.7721	-	7471A	PPB	17 May 2017 16:17:12	U	SMPL
ICB61 - 1	ICB61	-75	-0.0363	-	7471A	PPB	17 May 2017 16:19:27	U	SMPL
CCV26 - 1	CCV26	21188	4.9658	-	7471A	PPB	17 May 2017 16:21:43	U	SMPL
CCB26 - 1	CCB26	-73	-0.0358	-	7471A	PPB	17 May 2017 16:23:59	U	SMPL
CRA - 1	CRA	873	0.1867	-	7471A	PPB	17 May 2017 16:26:14	U	SMPL
HighStd - 1	HighStd	42045	9.8723	-	7471A	PPB	17 May 2017 16:28:30	U	SMPL
ChkStd - 1	ChkStd	28828	6.7631	-	7471A	PPB	17 May 2017 16:30:46	U	SMPL
PB99012BL - 1	PBS	-73	-0.0358	-	7471A	PPB	17 May 2017 16:33:03	U	SMPL
PB99012BS - 1	LCSS	13977	3.2694	-	7471A	PPB	17 May 2017 16:35:20	U	SMPL
I3172-01 - 1	SS-02	2484	0.5657	-	7471A	PPB	17 May 2017 16:37:37	U	SMPL
I3172-02 - 1	SS-03	2476	0.5638	-	7471A	PPB	17 May 2017 16:39:55	U	SMPL
I3172-03 - 1	SS-04	159	0.0187	-	7471A	PPB	17 May 2017 16:42:12	U	SMPL
I3177-01 - 1	NORTH-A	11397	2.6625	-	7471A	PPB	17 May 2017 16:44:30	U	SMPL
I3177-02 - 1	SOUTH-A	1363	0.3020	-	7471A	PPB	17 May 2017 16:46:46	U	SMPL
CCV27 - 1	CCV27	21341	5.0018	-	7471A	PPB	17 May 2017 16:49:02	U	SMPL
CCB27 - 1	CCB27	-54	-0.0314	-	7471A	PPB	17 May 2017 16:51:18	U	SMPL
I3177-03 - 1	EAST	4156	0.9590	-	7471A	PPB	17 May 2017 16:53:34	U	SMPL
I3177-04 - 1	FLOOR-A	2369	0.5386	-	7471A	PPB	17 May 2017 16:55:50	U	SMPL
I3177-04DUP - 1	FLOOR-ADUP	2014	0.4551	-	7471A	PPB	17 May 2017 16:58:06	U	SMPL
I3177-04MS - 1	FLOOR-AMS	17812	4.1716	-	7471A	PPB	17 May 2017 17:00:23	U	SMPL
I3177-04MSD - 1	FLOOR-AMSD	16820	3.9382	-	7471A	PPB	17 May 2017 17:02:40	U	SMPL
I3184-05 - 1	COMPOSITE1	2742	0.6264	-	7471A	PPB	17 May 2017 17:04:56	U	SMPL
I3184-06 - 1	COMPOSITE2	3382	0.7769	-	7471A	PPB	17 May 2017 17:07:14	U	SMPL
I3177-04LX5 - 1		475	0.0931	-	7471A	PPB	17 May 2017 17:09:31	U	SMPL
I3177-04A - 1		19430	4.5522	-	7471A	PPB	17 May 2017 17:11:49	U	SMPL
CCV28 - 1	CCV28	21435	5.0239	-	7471A	PPB	17 May 2017 17:14:06	U	SMPL
CCB28 - 1	CCB28	-57	-0.0321	-	7471A	PPB	17 May 2017 17:16:22	U	SMPL