



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87527

Temp.: 24.0 °C

BalanceID: WC SC-4

SLOPE.: 98.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/18/2017	09:25
2	CAL2	1	Water			7.02	05/18/2017	09:26
3	CAL3	1	Water			10.04	05/18/2017	09:27
4	ICV	1	Water			7.00	05/18/2017	09:30
5	CCV1	1	Water			2.02	05/18/2017	09:31
6	I3199-07	1	Solid	20.02	20	6.33	05/18/2017	09:32
7	I3199-07DUP	1	Solid	20.03	20	6.34	05/18/2017	09:33
8	I3199-08	1	Solid	20.04	20	6.83	05/18/2017	09:34
9	I3199-09	1	Solid	20.02	20	8.17	05/18/2017	09:35
10	I3202-15	1	Solid	20.01	20	7.26	05/18/2017	09:36
11	I3202-16	1	Solid	20.04	20	8.23	05/18/2017	09:37
12	I3202-17	1	Solid	20.02	20	7.77	05/18/2017	09:39
13	I3202-18	1	Solid	20.03	20	7.35	05/18/2017	09:41
14	I3202-19	1	Solid	20.00	20	7.38	05/18/2017	09:42
15	I3202-20	1	Solid	20.02	20	8.51	05/18/2017	09:43
16	CCV2	1	Water			12.02	05/18/2017	09:45
17	I3202-21	1	Solid	20.04	20	8.08	05/18/2017	09:46
18	I3202-22	1	Solid	20.02	20	6.63	05/18/2017	09:47
19	I3202-23	1	Solid	20.04	20	8.12	05/18/2017	09:49
20	I3202-24	1	Solid	20.01	20	10.50	05/18/2017	09:50

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3202-25	1	Solid	20.02	20	8.47	05/18/2017	09:51
22	I3207-04	1	Solid	20.04	20	5.67	05/18/2017	09:52
23	I3207-08	1	Solid	20.02	20	6.00	05/18/2017	09:53
24	I3207-08DUP	1	Solid	20.03	20	6.01	05/18/2017	09:54
25	CCV3	1	Water			12.01	05/18/2017	10:00

2887527

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3201

WorkList ID : 98579

Date : 5/18/2017 8:57:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3199-07	Corrosivity	Cool 4 deg C	LIRO01	H22	SB-04-COMP	05/15/2017	9045C
	Solid	I3199-08	Corrosivity	Cool 4 deg C	LIRO01	H22	SB-08-COMP	05/15/2017	9045C
	Solid	I3199-09	Corrosivity	Cool 4 deg C	LIRO01	H22	SB-11-COMP	05/15/2017	9045C
	Solid	I3202-15	Corrosivity	Cool 4 deg C	TULL02	H52	SB-1-COMP	05/16/2017	9045C
	Solid	I3202-16	Corrosivity	Cool 4 deg C	TULL02	H52	SB-2-COMP	05/16/2017	9045C
	Solid	I3202-17	Corrosivity	Cool 4 deg C	TULL02	H52	SB-3-COMP	05/16/2017	9045C
	Solid	I3202-18	Corrosivity	Cool 4 deg C	TULL02	H52	SB-4-COMP	05/16/2017	9045C
	Solid	I3202-19	Corrosivity	Cool 4 deg C	TULL02	H52	SB-5-COMP	05/16/2017	9045C
	Solid	I3202-20	Corrosivity	Cool 4 deg C	TULL02	H52	SB-6-COMP	05/16/2017	9045C
	Solid	I3202-21	Corrosivity	Cool 4 deg C	TULL02	H52	SB-7-COMP	05/16/2017	9045C
	Solid	I3202-22	Corrosivity	Cool 4 deg C	TULL02	H52	SB-8-COMP	05/16/2017	9045C
	Solid	I3202-23	Corrosivity	Cool 4 deg C	TULL02	H52	SB-9-COMP	05/16/2017	9045C
	Solid	I3202-24	Corrosivity	Cool 4 deg C	TULL02	H52	SB-10-COMP	05/16/2017	9045C
	Solid	I3202-25	Corrosivity	Cool 4 deg C	TULL02	H52	SB-11-COMP	05/16/2017	9045C
05/24/2017	Solid	I3207-04	Corrosivity	Cool 4 deg C	PSEG01	G61	TP-11-TP-12	05/17/2017	9045C
05/24/2017	Solid	I3207-08	Corrosivity	Cool 4 deg C	PSEG01	G61	TP-9-TP-10	05/17/2017	9045C

Date/Time 05-18-17 8:57 AM

Received by: SSRelinquished by: CS

Date/Time 05-18-17 8:57 AM

Received by: CSRelinquished by: JP