

LB 87528


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

35213



A= 0.0000e+000

B= 2.8407e-004

C= -1.2403e-002

Rho= 0.9998609

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.060	-0.060	-168	0.000	-168				
0.2	0.200	0.202	0.002	755	0.0 %	755				
2.5	2.500	2.531	0.031	8955	0.0 %	8955				
5.0	5.000	5.111	0.111	18036	0.0 %	18036				
7.5	7.500	7.425	-0.075	26180	0.0 %	26180				
10.0	10.000	9.991	-0.009	35213	0.0 %	35213				

LB87528 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 May 2017 09:24:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-168	-	0.0000	7470A	PPB	18 May 2017 09:31:53	S	Std
0.2 - 1	50.2	755	-	0.2000	7470A	PPB	18 May 2017 09:34:01	S	Std
2.5 - 1	52.5	8955	-	2.5000	7470A	PPB	18 May 2017 09:36:10	S	Std
5.0 - 1	55	18036	-	5.0000	7470A	PPB	18 May 2017 09:38:18	S	Std
7.5 - 1	57.5	26180	-	7.5000	7470A	PPB	18 May 2017 09:40:26	S	Std
10.0 - 1	50	35213	-	10.0000	7470A	PPB	18 May 2017 09:45:04	S	Std
ICV62 - 1	ICV62	14658	4.1516	-	7470A	PPB	18 May 2017 09:47:41	U	SMPL
ICB62 - 1	ICB62	-13	-0.0161	-	7470A	PPB	18 May 2017 09:49:48	U	SMPL
CCV29 - 1	CCV29	18218	5.1629	-	7470A	PPB	18 May 2017 09:51:54	U	SMPL
CCB29 - 1	CCB29	-158	-0.0573	-	7470A	PPB	18 May 2017 09:54:01	U	SMPL
CRA - 1	CRA	572	0.1501	-	7470A	PPB	18 May 2017 09:56:07	U	SMPL
HighStd - 1	HighStd	34960	9.9189	-	7470A	PPB	18 May 2017 09:58:14	U	SMPL
ChkStd - 1	ChkStd	25167	7.1369	-	7470A	PPB	18 May 2017 10:00:21	U	SMPL
PB99031BL - 1	PBW	-99	-0.0405	-	7470A	PPB	18 May 2017 10:02:29	U	SMPL
PB99031BS - 1	LCSW	14399	4.0780	-	7470A	PPB	18 May 2017 10:04:37	U	SMPL
PB99031TB - 1	PB99011TB	32	-0.0033	-	7470A	PPB	18 May 2017 10:06:45	U	SMPL
I3184-07 - 1	COMPOSITE1	-88	-0.0374	-	7470A	PPB	18 May 2017 10:08:53	U	SMPL
I3184-08 - 1	COMPOSITE2	-145	-0.0536	-	7470A	PPB	18 May 2017 10:11:01	U	SMPL
I3184-08DUP - 1	COMPOSITE2DUP	-91	-0.0383	-	7470A	PPB	18 May 2017 10:13:09	U	SMPL
I3184-08MS - 1	COMPOSITE2MS	13375	3.7871	-	7470A	PPB	18 May 2017 10:15:16	U	SMPL
CCV30 - 1	CCV30	17916	5.0771	-	7470A	PPB	18 May 2017 10:17:23	U	SMPL
CCB30 - 1	CCB30	-164	-0.0590	-	7470A	PPB	18 May 2017 10:19:29	U	SMPL
I3184-08MSD - 1	COMPOSITE2MSD	13553	3.8377	-	7470A	PPB	18 May 2017 10:21:35	U	SMPL
PB99032BL - 1	PBW	-112	-0.0442	-	7470A	PPB	18 May 2017 10:23:42	U	SMPL
PB99032BS - 1	LCSW	13189	3.7343	-	7470A	PPB	18 May 2017 10:25:48	U	SMPL
I3181-01 - 1	WELLA	-163	-0.0587	-	7470A	PPB	18 May 2017 10:27:55	U	SMPL
I3181-02 - 1	WELLB	-120	-0.0465	-	7470A	PPB	18 May 2017 10:30:03	U	SMPL
I3181-02DUP - 1	WELLBDUP	-104	-0.0419	-	7470A	PPB	18 May 2017 10:32:10	U	SMPL
I3181-02MS - 1	WELLBMS	12147	3.4383	-	7470A	PPB	18 May 2017 10:34:19	U	SMPL
I3181-02MSD - 1	WELLBMSD	12107	3.4269	-	7470A	PPB	18 May 2017 10:36:27	U	SMPL
I3181-03 - 1	CITYC	-139	-0.0519	-	7470A	PPB	18 May 2017 10:38:35	U	SMPL
I3181-04 - 1	CITYD	-90	-0.0380	-	7470A	PPB	18 May 2017 10:40:42	U	SMPL
CCV31 - 1	CCV31	18127	5.1370	-	7470A	PPB	18 May 2017 10:42:49	U	SMPL
CCB31 - 1	CCB31	-138	-0.0516	-	7470A	PPB	18 May 2017 10:44:56	U	SMPL
I3184-08LX5 - 1		-53	-0.0275	-	7470A	PPB	18 May 2017 10:47:03	U	SMPL
I3184-08A - 1		14436	4.0885	-	7470A	PPB	18 May 2017 10:49:10	U	SMPL
I3181-02LX5 - 1		-86	-0.0368	-	7470A	PPB	18 May 2017 10:51:17	U	SMPL
I3181-02A - 1		14715	4.1678	-	7470A	PPB	18 May 2017 10:53:24	U	SMPL
CCV32 - 1	CCV32	18121	5.1353	-	7470A	PPB	18 May 2017 10:55:32	U	SMPL
CCB32 - 1	CCB32	-88	-0.0374	-	7470A	PPB	18 May 2017 10:57:39	U	SMPL