



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB87538

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	05/18/2017	11:20
2	I3150-01		1	>100.0	>212.0	05/18/2017	11:50
3	I3150-01DUP		1	>100.0	>212.0	05/18/2017	12:15
4	I3150-02		1	>100.0	>212.0	05/18/2017	12:45
5	I3171-01		1	>100.0	>212.0	05/18/2017	13:20
6	I3171-02		1	>100.0	>212.0	05/18/2017	13:50

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 87538

Worklist Name : FP51817

Worklist ID : 98588

Date : 5/18/2017 10:38:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	13150-01	Flash Point	Cool 4 deg C	TULL02	H61	WS-1	05/12/2017	1010A
	Water	13150-02	Flash Point	Cool 4 deg C	TULL02	H61	WS-2	05/12/2017	1010A
	Water	13171-01	Flash Point	Cool 4 deg C	TULL02	H32	WS-4	05/15/2017	1010A
	Water	13171-02	Flash Point	Cool 4 deg C	TULL02	H32	WS-5	05/15/2017	1010A

Date/Time 5/18/17 10:45 AM
 Received by: KS
 Relinquished by: JR

Date/Time 5/18/17 14:00
 Received by: JR
 Relinquished by: KS