



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87543

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/18/2017	18:30
2	CAL2	1	Water			7.01	05/18/2017	18:31
3	CAL3	1	Water			10.03	05/18/2017	18:32
4	ICV	1	Water			7.00	05/18/2017	18:33
5	CCV1	1	Water			2.01	05/18/2017	18:35
6	I3125-05	1	Solid	20.02	20	7.61	05/18/2017	18:36
7	I3125-05DUP	1	Solid	20.04	20	7.62	05/18/2017	18:37
8	I3227-03	1	Solid	20.02	20	5.30	05/18/2017	18:39
9	I3228-03	1	Solid	20.01	20	8.58	05/18/2017	18:40
10	I3235-03	1	Solid	20.03	20	8.18	05/18/2017	18:41
11	CCV2	1	Water			12.01	05/18/2017	18:42

2987543

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3235

WorkList ID : 98611

Date : 5/18/2017 6:22:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/17/2017	Solid	I3125-05	Corrosivity	Cool 4 deg C	ENTA04	H32	WC-AOC-17-B5-COMP	05/18/2017	9045C
	Solid	I3227-03	Corrosivity	Cool 4 deg C	PSEG01	F21	RO-11939	05/18/2017	9045C
	Solid	I3228-03	Corrosivity	Cool 4 deg C	PSEG01	F21	MH-9	05/18/2017	9045C
	Solid	I3235-03	Corrosivity	Cool 4 deg C	PSEG01	F31	TW1-TW2	05/18/2017	9045C

Date/Time 05-18-17 6:22:17
Received by: SP
Relinquished by: CP

Date/Time 05-18-17 7:30:12
Received by: SP
Relinquished by: SP