



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87544

Temp.: 23.2 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/18/2017	18:30
2	CAL2	1	Water			7.01	05/18/2017	18:31
3	CAL3	1	Water			10.03	05/18/2017	18:32
4	ICV	1	Water			7.00	05/19/2017	18:33
5	CCV1	1	Water			2.01	05/18/2017	18:40
6	I3235-07	1	Water			6.57	05/18/2017	18:41
7	I3235-07DUP	1	Water			6.58	05/18/2017	18:42
8	CCV2	1	Water			12.01	05/18/2017	18:43

WORKLIST(Hardcopy Internal Chain)

20 87544

WorkList Name : CORROSIVITYW13235 WorkList ID : 98612 Date : 5/18/2017 6:23:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3235-07	Corrosivity	Cool 4 deg C	PSEG01	F31	TW1-TW2	05/18/2017	9040C

Date/Time 05-18-17 6:22:11m
 Received by: SP
 Relinquished by: CR

Date/Time 05-18-17 7:30:00m
 Received by: CR
 Relinquished by: SP