



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87546

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/18/2017	18:30
2	CAL2	1	Water			7.01	05/18/2017	18:31
3	CAL3	1	Water			10.03	05/18/2017	18:32
4	ICV	1	Water			7.00	05/18/2017	18:33
5	CCV1	1	Water			2.02	05/18/2017	18:45
6	I3232-01	1	Solid	20.03	20	6.20	05/18/2017	18:46
7	I3232-01DUP	1	Solid	20.05	20	6.21	05/18/2017	18:47
8	I3236-01	1	Solid	20.02	20	5.83	05/18/2017	18:49
9	CCV2	1	Water			12.02	05/18/2017	18:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3232

WorkList ID : 98614

Date : 5/18/2017 6:25:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3232-01	pH	Cool 4 deg C	TULL02	I12	FK-PS-01-003	05/18/2017	9045C
	Solid	I3236-01	pH	Cool 4 deg C	TULL02	I32	FK-PS-01_003	05/18/2017	9045C

Date/Time 05-18-17 6:22 PM
 Received by: SR
 Relinquished by: CP

Date/Time 05-18-17 7:30 PM
 Received by: CP
 Relinquished by: JD