


7470A

LB 87552
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36069



A= 0.0000e+000

B= 2.7697e-004

C= 2.2246e-002

Rho= 0.9998756

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	13	0.000	13				
0.2	0.200	0.196	-0.004	628	0.0 %	628				
2.5	2.500	2.426	-0.074	8680	0.0 %	8680				
5.0	5.000	5.101	0.101	18338	0.0 %	18338				
7.5	7.500	7.438	-0.062	26776	0.0 %	26776				
10.0	10.000	10.012	0.012	36069	0.0 %	36069				

LB87552 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 May 2017 09:27:34

Sample ID	Extended ID	μ Abs.	Conc.	Std. ConcMethod	Units	Date	Type	Type
0.0 - 1	50	13	-	0.00007470A	PPB	19 May 2017 09:32:27	S	Std
0.2 - 1	50.2	628	-	0.20007470A	PPB	19 May 2017 09:34:35	S	Std
2.5 - 1	52.5	8680	-	2.50007470A	PPB	19 May 2017 09:36:44	S	Std
5.0 - 1	55	18338	-	5.00007470A	PPB	19 May 2017 09:38:52	S	Std
7.5 - 1	57.5	26776	-	7.50007470A	PPB	19 May 2017 09:41:00	S	Std
10.0 - 1	510	36069	-	10.00007470A	PPB	19 May 2017 09:43:07	S	Std
ICV64 - 1	ICV64	13936	3.8820	-7470A	PPB	19 May 2017 09:45:47	U	SMPL
ICB64 - 1	ICB64	-208	-0.0354	-7470A	PPB	19 May 2017 09:47:53	U	SMPL
CCV44 - 1	CCV44	17685	4.9204	-7470A	PPB	19 May 2017 09:50:00	U	SMPL
CCB44 - 1	CCB44	-270	-0.0525	-7470A	PPB	19 May 2017 09:52:06	U	SMPL
CRA - 1	CRA	648	0.2017	-7470A	PPB	19 May 2017 09:54:13	U	SMPL
HighStd - 1	HighStd	35003	9.7169	-7470A	PPB	19 May 2017 09:56:19	U	SMPL
ChkStd - 1	ChkStd	25092	6.9719	-7470A	PPB	19 May 2017 09:58:25	U	SMPL
PB99074BL - 1	PBW	-48	0.0090	-7470A	PPB	19 May 2017 10:00:33	U	SMPL
PB99074BS - 1	LCSW	13104	3.6516	-7470A	PPB	19 May 2017 10:02:41	U	SMPL
PB99074TB - 1	PB99019TB	-273	-0.0534	-7470A	PPB	19 May 2017 10:04:48	U	SMPL
I3182-05 - 1	VNJ-204-228-ETGI-275	-162	-0.0226	-7470A	PPB	19 May 2017 10:06:56	U	SMPL
I3182-11 - 1	ETGI-137-339-120	-15	0.0181	-7470A	PPB	19 May 2017 10:09:05	U	SMPL
I3182-17 - 1	72-11932-72-11930-ETGI-338	61	0.0391	-7470A	PPB	19 May 2017 10:11:13	U	SMPL
I3185-01 - 1	OWS-2	896	0.2704	-7470A	PPB	19 May 2017 10:13:19	U	SMPL
CCV45 - 1	CCV45	17983	5.0029	-7470A	PPB	19 May 2017 10:15:26	U	SMPL
CCB45 - 1	CCB45	-263	-0.0506	-7470A	PPB	19 May 2017 10:17:33	U	SMPL
I3187-05 - 1	SP-COMP	-208	-0.0354	-7470A	PPB	19 May 2017 10:19:40	U	SMPL
I3188-03 - 1	RO-120-20	-113	-0.0091	-7470A	PPB	19 May 2017 10:21:47	U	SMPL
I3189-05 - 1	TP-15	-112	-0.0088	-7470A	PPB	19 May 2017 10:23:54	U	SMPL
I3189-10 - 1	TP-13-14	-78	0.0006	-7470A	PPB	19 May 2017 10:26:01	U	SMPL
I3199-07 - 1	SB-04-COMP	-130	-0.0138	-7470A	PPB	19 May 2017 10:28:09	U	SMPL
I3199-08 - 1	SB-08-COMP	-67	0.0037	-7470A	PPB	19 May 2017 10:30:16	U	SMPL
I3199-09 - 1	SB-11-COMP	-67	0.0037	-7470A	PPB	19 May 2017 10:32:25	U	SMPL
I3199-09DUP - 1	SB-11-COMPDUP	-77	0.0009	-7470A	PPB	19 May 2017 10:34:33	U	SMPL
I3199-09MS - 1	SB-11-COMPMS	12703	3.5405	-7470A	PPB	19 May 2017 10:36:42	U	SMPL
I3199-09MSD - 1	SB-11-COMPMSD	12714	3.5436	-7470A	PPB	19 May 2017 10:38:49	U	SMPL
CCV46 - 1	CCV46	18355	5.1060	-7470A	PPB	19 May 2017 10:40:55	U	SMPL
CCB46 - 1	CCB46	-150	-0.0193	-7470A	PPB	19 May 2017 10:43:02	U	SMPL
I3199-09LX5 - 1		-86	-0.0016	-7470A	PPB	19 May 2017 10:45:09	U	SMPL
I3199-09A - 1		13937	3.8823	-7470A	PPB	19 May 2017 10:47:16	U	SMPL
CCV47 - 1	CCV47	18244	5.0752	-7470A	PPB	19 May 2017 10:49:23	U	SMPL
CCB47 - 1	CCB47	-162	-0.0226	-7470A	PPB	19 May 2017 10:51:30	U	SMPL