



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87564

Temp.: 22.1 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/19/2017	09:35
2	CAL2	1	Water			7.00	05/19/2017	09:36
3	CAL3	1	Water			10.02	05/19/2017	09:37
4	ICV	1	Water			6.99	05/19/2017	09:38
5	CCV1	1	Water			2.02	05/19/2017	09:40
6	I3245-01	1	Water			6.99	05/19/2017	09:41
7	I3245-02	1	Water			7.11	05/19/2017	09:42
8	I3245-03	1	Water			6.52	05/19/2017	09:43
9	I3245-03DUP	1	Water			6.53	05/19/2017	09:45
10	CCV2	1	Water			12.01	05/19/2017	09:46

WORKLIST(Hardcopy Internal Chain)

B 2887564

WorkList Name : ph i3245

WorkList ID : 98669

Date : 5/19/2017 9:25:14 am

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3245-01 - E	pH	Cool 4 deg C	WOOD06	I61	GMW-7	05/18/2017	SM 4500-PH B
	Water	I3245-02 - E	pH	Cool 4 deg C	WOOD06	I61	GMW-X	05/18/2017	SM 4500-PH B
	Water	I3245-03 E	pH	Cool 4 deg C	WOOD06	I61	GMW-10	05/18/2017	SM 4500-PH B

Date/Time 05.19.17 9:25 AM

Received by: JH

Relinquished by: CR

Date/Time 05.19.17 10:30 AM

Received by: CR

Relinquished by: CR