

LB87570


7471A

INSTRUMENT ID: CV1

Linear

μ Abs.:

45121



A= 0.0000e+000

B= 2.2090e-004

C= -3.3042e-003

Rho= 0.9999302

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	98	0.000	98				
0.2	0.200	0.215	0.015	990	0.0 %	990				
2.5	2.500	2.424	-0.076	10986	0.0 %	10986				
5.0	5.000	5.022	0.022	22747	0.0 %	22747				
7.5	7.500	7.557	0.057	34224	0.0 %	34224				
10.0	10.000	9.964	-0.036	45121	0.0 %	45121				

LB87570 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 19 May 2017 14:45:21

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	98	-	0.00007471A	PPB	19 May 2017 14:48:19	S	Std
0.2 - 1	50.2	990	-	0.20007471A	PPB	19 May 2017 14:50:36	S	Std
2.5 - 1	52.5	10986	-	2.50007471A	PPB	19 May 2017 14:52:54	S	Std
5.0 - 1	55	22747	-	5.00007471A	PPB	19 May 2017 14:55:11	S	Std
7.5 - 1	57.5	34224	-	7.50007471A	PPB	19 May 2017 14:57:28	S	Std
10.0 - 1	50	45121	-	10.00007471A	PPB	19 May 2017 14:59:45	S	Std
ICV65 - 1	ICV65	17247	3.8066	-7471A	PPB	19 May 2017 15:02:38	U	SMPL
ICB65 - 1	ICB65	16	0.0002	-7471A	PPB	19 May 2017 15:04:53	U	SMPL
CCV50 - 1	CCV50	22564	4.9812	-7471A	PPB	19 May 2017 15:07:09	U	SMPL
CCB50 - 1	CCB50	21	0.0013	-7471A	PPB	19 May 2017 15:09:25	U	SMPL
CRA - 1	CRA	956	0.2079	-7471A	PPB	19 May 2017 15:11:41	U	SMPL
HighStd - 1	HighStd	44278	9.7779	-7471A	PPB	19 May 2017 15:13:56	U	SMPL
ChkStd - 1	ChkStd	31252	6.9004	-7471A	PPB	19 May 2017 15:16:13	U	SMPL
PB99077BL - 1	PBS	-145	-0.0353	-7471A	PPB	19 May 2017 15:18:29	U	SMPL
PB99077BS - 1	LCSS	15465	3.4130	-7471A	PPB	19 May 2017 15:20:46	U	SMPL
I3200-01 - 1	POST-EX-34-20170517	1790	0.3921	-7471A	PPB	19 May 2017 15:23:03	U	SMPL
I3200-02 - 1	POST-EX-35-20170517	19858	4.3834	-7471A	PPB	19 May 2017 15:25:20	U	SMPL
I3200-03 - 1	POST-EX-36-20170517	18312	4.0419	-7471A	PPB	19 May 2017 15:27:37	U	SMPL
I3200-04 - 1	POST-EX-DUP2-20170517	13636	3.0090	-7471A	PPB	19 May 2017 15:29:55	U	SMPL
I3200-05 - 1	POST-EX-37-20170517	23296	5.1429	-7471A	PPB	19 May 2017 15:32:11	U	SMPL
CCV51 - 1	CCV51	22557	4.9796	-7471A	PPB	19 May 2017 15:34:27	U	SMPL
CCB51 - 1	CCB51	33	0.0040	-7471A	PPB	19 May 2017 15:36:43	U	SMPL
I3200-06 - 1	POST-EX-38-20170517	22221	4.9054	-7471A	PPB	19 May 2017 15:39:00	U	SMPL
I3200-06DUP - 1	POST-EX-38-20170517DUP	20324	4.4864	-7471A	PPB	19 May 2017 15:41:16	U	SMPL
I3200-07 - 1	I3200-06MS	39266	8.6707	-7471A	PPB	19 May 2017 15:43:32	U	SMPL
I3200-08 - 1	I3200-06MSD	39156	8.6464	-7471A	PPB	19 May 2017 15:48:22	U	SMPL
I3200-09 - 1	POST-EX-39-20170517	4067	0.8951	-7471A	PPB	19 May 2017 15:50:38	U	SMPL
I3200-10 - 1	POST-EX-40-20170517	17769	3.9220	-7471A	PPB	19 May 2017 15:52:55	U	SMPL
I3200-11 - 1	POST-EX-41-20170517	32327	7.1379	-7471A	PPB	19 May 2017 15:55:12	U	SMPL
I3202-01 - 1	SB-1COMP	1074	0.2339	-7471A	PPB	19 May 2017 15:57:30	U	SMPL
I3202-02 - 1	SB-2COMP	292	0.0612	-7471A	PPB	19 May 2017 15:59:48	U	SMPL
I3202-03 - 1	SB-3COMP	193	0.0393	-7471A	PPB	19 May 2017 16:02:04	U	SMPL
CCV52 - 1	CCV52	22677	5.0062	-7471A	PPB	19 May 2017 16:04:21	U	SMPL
CCB52 - 1	CCB52	61	0.0102	-7471A	PPB	19 May 2017 16:06:37	U	SMPL
I3202-04 - 1	SB-4COMP	556	0.1195	-7471A	PPB	19 May 2017 16:08:53	U	SMPL
I3202-05 - 1	SB-5COMP	262	0.0546	-7471A	PPB	19 May 2017 16:11:10	U	SMPL
I3202-06 - 1	SB-6COMP	307	0.0645	-7471A	PPB	19 May 2017 16:13:26	U	SMPL
I3202-07 - 1	SB-7COMP	680	0.1469	-7471A	PPB	19 May 2017 16:15:43	U	SMPL
I3202-08 - 1	SB-8COMP	967	0.2103	-7471A	PPB	19 May 2017 16:17:59	U	SMPL
I3202-09 - 1	SB-9COMP	176	0.0356	-7471A	PPB	19 May 2017 16:20:16	U	SMPL
I3202-10 - 1	SB-10COMP	2133	0.4679	-7471A	PPB	19 May 2017 16:22:34	U	SMPL
I3202-11 - 1	SB-11COMP	174	0.0351	-7471A	PPB	19 May 2017 16:24:51	U	SMPL
PB99078BL - 1	PBS	-53	-0.0150	-7471A	PPB	19 May 2017 16:27:09	U	SMPL
PB99078BS - 1	LCSS	15942	3.5184	-7471A	PPB	19 May 2017 16:29:26	U	SMPL
CCV53 - 1	CCV53	22854	5.0453	-7471A	PPB	19 May 2017 16:31:42	U	SMPL
CCB53 - 1	CCB53	92	0.0170	-7471A	PPB	19 May 2017 16:33:58	U	SMPL
I3203-01 - 1	E2-CONCRETE-BATCH-10	6013	1.3250	-7471A	PPB	19 May 2017 16:36:14	U	SMPL
I3203-02 - 1	E2-CONCRETE-BATCH-11-BULK	10209	2.2519	-7471A	PPB	19 May 2017 16:38:31	U	SMPL
I3203-03 - 1	E2-R4-BASE	16587	3.6608	-7471A	PPB	19 May 2017 16:40:47	U	SMPL
I3203-03DUP - 1	E2-R4-BASEDUP	17938	3.9593	-7471A	PPB	19 May 2017 16:43:03	U	SMPL
I3203-03MS - 1	E2-R4-BASEMS	35855	7.9172	-7471A	PPB	19 May 2017 16:50:20	U	SMPL
I3203-03MSD - 1	E2-R4-BASEMSD	34858	7.6970	-7471A	PPB	19 May 2017 16:52:37	U	SMPL
I3203-04 - 1	E2-S5-BASE	58226	12.8591	-7471A	PPB	19 May 2017 16:54:54	U	SMPL
I3203-05 - 1	E2-T7-BASE	64130	14.1633	-7471A	PPB	19 May 2017 16:57:11	U	SMPL
I3203-06 - 1	E2-T8-BASE	1406	0.3073	-7471A	PPB	19 May 2017 16:59:29	U	SMPL
I3207-01 - 1	TP-11-TP-12	389	0.0826	-7471A	PPB	19 May 2017 17:01:46	U	SMPL
CCV54 - 1	CCV54	21740	4.7992	-7471A	PPB	19 May 2017 17:04:03	U	SMPL
CCB54 - 1	CCB54	87	0.0159	-7471A	PPB	19 May 2017 17:06:19	U	SMPL
I3207-05 - 1	TP-9-TP-10	102	0.0192	-7471A	PPB	19 May 2017 17:08:35	U	SMPL
I3210-01 - 1	SS-1	2503	0.5496	-7471A	PPB	19 May 2017 17:10:52	U	SMPL
I3200-06LX5 - 1		3987	0.8774	-7471A	PPB	19 May 2017 17:13:08	U	SMPL
I3200-06A - 1		37570	8.2961	-7471A	PPB	19 May 2017 17:15:24	U	SMPL
I3203-03LX5 - 1		3042	0.6687	-7471A	PPB	19 May 2017 17:17:41	U	SMPL
I3203-03A - 1		32139	7.0964	-7471A	PPB	19 May 2017 17:19:58	U	SMPL
I3203-04DLX2 - 1		27989	6.1796	-7471A	PPB	19 May 2017 17:22:16	U	SMPL
I3203-05DLX2 - 1		32141	7.0968	-7471A	PPB	19 May 2017 17:24:33	U	SMPL
CCV55 - 1	CCV55	21951	4.8458	-7471A	PPB	19 May 2017 17:26:51	U	SMPL
CCB55 - 1	CCB55	92	0.0170	-7471A	PPB	19 May 2017 17:29:08	U	SMPL