



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87576

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/19/2017	18:30
2	CAL2	1	Water			7.01	05/19/2017	18:31
3	CAL3	1	Water			10.03	05/19/2017	18:32
4	ICV	1	Water			7.00	05/19/2017	18:33
5	CCV1	1	Water			2.02	05/19/2017	18:35
6	I3243-05	1	Solid	20.02	20	5.56	05/19/2017	18:36
7	I3243-05DUP	1	Solid	20.04	20	5.57	05/19/2017	18:37
8	I3243-10	1	Solid	20.01	20	5.51	05/20/2017	18:39
9	I3249-18	1	Solid	20.02	20	7.77	05/19/2017	18:40
10	I3249-20	1	Solid	20.04	20	7.84	05/19/2017	18:41
11	I3249-22	1	Solid	20.03	20	7.87	05/19/2017	18:42
12	I3249-24	1	Solid	20.01	20	8.64	05/19/2017	18:43
13	I3249-26	1	Solid	20.05	20	7.76	05/19/2017	18:45
14	I3249-28	1	Solid	20.03	20	7.25	05/19/2017	18:46
15	I3249-30	1	Solid	20.05	20	7.61	05/19/2017	18:47
16	CCV2	1	Water			12.01	05/19/2017	18:49
17	I3249-32	1	Solid	20.04	20	8.10	05/19/2017	18:50
18	I3250-02	1	Solid	20.03	20	9.17	05/19/2017	18:51
19	I3278-03	1	Solid	20.01	20	6.87	05/19/2017	18:52
20	I3278-06	1	Solid	20.02	20	7.54	05/19/2017	18:53

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3278-09	1	Solid	20.03	20	8.72	05/19/2017	18:55
22	I3278-09DUP	1	Solid	20.04	20	8.73	05/19/2017	18:56
23	CCV3	1	Water			12.02	05/19/2017	19:00

WORKLIST(Hardcopy Internal Chain)

18 87376

WorkList Name : CORROSIONITY I3249

WorkList ID : 98681

Date : 5/19/2017 7:24:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3243-05	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-7-TP-8	05/18/2017	9045C
	Solid	I3243-10	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-5-TP-6	05/18/2017	9045C
	Solid	I3249-18	Corrosivity	Cool 4 deg C	TULL02	I61	SB-12COMP	05/19/2017	9045C
	Solid	I3249-20	Corrosivity	Cool 4 deg C	TULL02	I61	SB-13COMP	05/19/2017	9045C
	Solid	I3249-22	Corrosivity	Cool 4 deg C	TULL02	I61	SB-14COMP	05/19/2017	9045C
	Solid	I3249-24	Corrosivity	Cool 4 deg C	TULL02	I61	SB-19COMP	05/19/2017	9045C
	Solid	I3249-26	Corrosivity	Cool 4 deg C	TULL02	I61	SB-18COMP	05/19/2017	9045C
	Solid	I3249-28	Corrosivity	Cool 4 deg C	TULL02	I61	SB-17COMP	05/19/2017	9045C
	Solid	I3249-30	Corrosivity	Cool 4 deg C	TULL02	I61	SB-16COMP	05/19/2017	9045C
	Solid	I3249-32	Corrosivity	Cool 4 deg C	TULL02	I61	SB-20COMP	05/19/2017	9045C
	Solid	I3250-02	Corrosivity	Cool 4 deg C	TULL02	I61	W.P.RD-C.AVE(SP)	05/18/2017	9045C
	Solid	I3278-03	Corrosivity	Cool 4 deg C	PSEG01	I61	RT-3873	05/19/2017	9045C
	Solid	I3278-06	Corrosivity	Cool 4 deg C	PSEG01	I61	RT-2669	05/19/2017	9045C
	Solid	I3278-09	Corrosivity	Cool 4 deg C	PSEG01	I61	RT-4632	05/19/2017	9045C

Date/Time 05-19-17

Received by: JI

Relinquished by:

Date/Time 05-19-17

Received by: JI

Relinquished by: JI