



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87580

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/20/2017	09:00
2	CAL2	1	Water			7.01	05/20/2017	09:01
3	CAL3	1	Water			10.02	05/20/2017	09:03
4	ICV	1	Water			7.00	05/20/2017	09:05
5	CCV1	1	Water			2.01	05/20/2017	09:06
6	I3252-01	1	Solid	20.02	20	3.61	05/20/2017	09:09
7	I3252-01DUP	1	Solid	20.03	20	3.62	05/20/2017	09:10
8	I3257-01	1	Solid	20.05	20	7.75	05/20/2017	09:11
9	I3261-01	1	Solid	20.01	20	5.60	05/20/2017	09:12
10	I3264-01	1	Solid	20.02	20	4.13	05/20/2017	09:13
11	CCV2	1	Water			12.01	05/20/2017	09:15

LB 84580

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3252

WorkList ID : 98697

Date : 5/20/2017 8:42:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3252-01	pH	Cool 4 deg C	TULL02	I32	NCC-BUCK-CLAY	05/18/2017	9045C
	Solid	I3257-01	pH	Cool 4 deg C	TULL02	I22	NCC-JM-COMPOST	05/18/2017	9045C
	Solid	I3261-01	pH	Cool 4 deg C	TULL02	I63	NCC-BUCK-3-1	05/18/2017	9045C
	Solid	I3264-01	pH	Cool 4 deg C	TULL02	I32	NCC-BUCK-FILL	05/18/2017	9045C

Date/Time 05/20/17 8:42:01 AM
Received by: JR
Relinquished by: JR

Date/Time 05/20/17 11:30 AM
Received by: JR
Relinquished by: JD