



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87581

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/20/2017	14:10
2	CAL2	1	Water			7.00	05/20/2017	14:11
3	CAL3	1	Water			10.04	05/20/2017	14:12
4	ICV	1	Water			7.01	05/20/2017	14:15
5	CCV1	1	Water			2.01	05/20/2017	14:16
6	I3214-18	1	Solid	20.02	20	8.17	05/20/2017	14:18
7	I3214-18DUP	1	Solid	20.04	20	8.18	05/20/2017	14:20
8	I3214-19	1	Solid	20.01	20	8.10	05/20/2017	14:21
9	I3214-20	1	Solid	20.02	20	8.18	05/20/2017	14:22
10	I3214-21	1	Solid	20.05	20	7.47	05/20/2017	14:25
11	I3214-22	1	Solid	20.01	20	7.18	05/20/2017	14:26
12	I3214-23	1	Solid	20.02	20	5.60	05/20/2017	14:27
13	I3215-21	1	Solid	20.03	20	8.16	05/20/2017	14:30
14	I3215-22	1	Solid	20.02	20	7.41	05/20/2017	14:31
15	I3215-23	1	Solid	20.01	20	9.36	05/20/2017	14:32
16	CCV2	1	Water			12.02	05/20/2017	14:33
17	I3215-24	1	Solid	20.03	20	8.62	05/20/2017	14:25
18	I3215-25	1	Solid	20.00	20	8.50	05/20/2017	14:26
19	I3215-26	1	Solid	20.03	20	7.68	05/20/2017	14:27
20	I3215-27	1	Solid	20.01	20	8.25	05/20/2017	14:30

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			12.01	05/20/2017	14:31

WORKLIST(Hardcopy Internal Chain)

87581

WorkList Name : CORROSIONITY I3214

WorkList ID : 98705

Date : 5/20/2017 12:26:37 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3214-18	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP2-WC8(0-8)	05/16/2017	9045C
	Solid	I3214-19	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP2-WC7(0-6.5)	05/16/2017	9045C
	Solid	I3214-20	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP2-WC6(0-6.5)	05/16/2017	9045C
	Solid	I3214-21	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP2-WC5(0-10)	05/16/2017	9045C
	Solid	I3214-22	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP2-WC5(10-24)	05/16/2017	9045C
	Solid	I3214-23	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP2-WC1(6-24)	05/16/2017	9045C
	Solid	I3215-21	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC11(0-2)	05/15/2017	9045C
	Solid	I3215-22	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC11(6-8)	05/15/2017	9045C
	Solid	I3215-23	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC9(0-9.8)	05/15/2017	9045C
	Solid	I3215-24	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC9(9.8-12)	05/15/2017	9045C
	Solid	I3215-25	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC7(0-4)	05/15/2017	9045C
	Solid	I3215-26	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC4(0-9.5)	05/15/2017	9045C
	Solid	I3215-27	Corrosivity	Cool 4 deg C	CHAZ02	I52	EP1-WC2(0-6)	05/15/2017	9045C

Date/Time 05-20-17 12:26 PM
 Received by: JP
 Relinquished by: JP

Date/Time 05-20-17 5:30 PM
 Received by: JP
 Relinquished by: JP