



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87582

Temp.: 19.6 °C

SLOPE.: 98.2

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2183         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2154         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W2140         |
| pH 2.00 Buffer                 | W2079         |
| Buffer Solution, PH12 (500ml)  | W2099         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 4.01        | 05/20/2017 | 12:35     |
| 2   | CAL2        | 1  | Water  |             |             | 7.01        | 05/20/2017 | 12:36     |
| 3   | CAL3        | 1  | Water  |             |             | 10.03       | 05/20/2017 | 12:37     |
| 4   | ICV         | 1  | Water  |             |             | 6.99        | 05/20/2017 | 12:39     |
| 5   | CCV1        | 1  | Water  |             |             | 2.01        | 05/19/2017 | 12:40     |
| 6   | I3239-01    | 1  | Water  |             |             | 10.77       | 05/19/2017 | 12:41     |
| 7   | I3239-01DUP | 1  | Water  |             |             | 10.78       | 05/20/2017 | 12:42     |
| 8   | I3239-02    | 1  | Water  |             |             | 11.99       | 05/20/2017 | 12:44     |
| 9   | CCV2        | 1  | Water  |             |             | 12.01       | 05/20/2017 | 12:45     |

2887582

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3286      WorkList ID : 98706      Date : 5/20/2017 12:27:25 PM

| Due Date | Matrix | Sample      | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|----------|--------|-------------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
|          | Water  | I3239-01 -1 | Corrosivity | Cool 4 deg C | SEAC01   | I32              | PURGE-WATER-A   | 05/19/2017   | 9040C  |
|          | Water  | I3239-02 -1 | Corrosivity | Cool 4 deg C | SEAC01   | I32              | PURGE-WATER-B   | 05/19/2017   | 9040C  |

Date/Time 05-20-17 12:27 PM  
 Received by: SR  
 Relinquished by: SR

Date/Time 05-20-17 2:00 PM  
 Received by: SR  
 Relinquished by: SR