


7470A

LB87594
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36683



A= 0.0000e+000

B= 2.7418e-004

C= -2.5465e-002

Rho= 0.9999551

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	160	0.000	160				
0.2	0.200	0.183	-0.017	759	0.0 %	759				
2.5	2.500	2.544	0.044	9372	0.0 %	9372				
5.0	5.000	4.942	-0.058	18119	0.0 %	18119				
7.5	7.500	7.480	-0.020	27375	0.0 %	27375				
10.0	10.000	10.032	0.032	36683	0.0 %	36683				

LB87594 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 May 2017 09:03:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	160	-	0.00007470A	PPB		22 May 2017 09:31:50	S	Std
0.2 - 1	50.2	759	-	0.20007470A	PPB		22 May 2017 09:33:58	S	Std
2.5 - 1	52.5	9372	-	2.50007470A	PPB		22 May 2017 09:36:06	S	Std
5.0 - 1	55	18119	-	5.00007470A	PPB		22 May 2017 09:38:14	S	Std
7.5 - 1	57.5	27375	-	7.50007470A	PPB		22 May 2017 09:40:22	S	Std
10.0 - 1	510	36683	-	10.00007470A	PPB		22 May 2017 09:44:59	S	Std
ICV66 - 1	ICV66	14461	3.9394	-7470A	PPB		22 May 2017 09:47:32	U	SMPL
ICB66 - 1	ICB66	208	0.0316	-7470A	PPB		22 May 2017 09:49:38	U	SMPL
CCV56 - 1	CCV56	18082	4.9322	-7470A	PPB		22 May 2017 09:51:44	U	SMPL
CCB56 - 1	CCB56	203	0.0302	-7470A	PPB		22 May 2017 09:53:51	U	SMPL
CRA - 1	CRA	865	0.2117	-7470A	PPB		22 May 2017 09:55:57	U	SMPL
HighStd - 1	HighStd	37294	10.1998	-7470A	PPB		22 May 2017 09:58:04	U	SMPL
ChkStd - 1	ChkStd	25830	7.0566	-7470A	PPB		22 May 2017 10:00:11	U	SMPL
PB99119BL - 1	PBW	75	-0.0049	-7470A	PPB		22 May 2017 10:02:19	U	SMPL
PB99119BS - 1	LCSW	14338	3.9057	-7470A	PPB		22 May 2017 10:04:26	U	SMPL
I3200-12 - 1	FB-20170517	-100	-0.0529	-7470A	PPB		22 May 2017 10:06:34	U	SMPL
I3235-04 - 1	TW1-TW2	2558	0.6759	-7470A	PPB		22 May 2017 10:08:42	U	SMPL
I3240-15 - 1	EP2-FB01	87	-0.0016	-7470A	PPB		22 May 2017 10:10:51	U	SMPL
I3241-11 - 1	EP1-FB01	115	0.0061	-7470A	PPB		22 May 2017 10:12:59	U	SMPL
I3242-01 - 1	VL-WWTP-003	920	0.2268	-7470A	PPB		22 May 2017 10:15:06	U	SMPL
CCV57 - 1	CCV57	17152	4.6773	-7470A	PPB		22 May 2017 10:17:13	U	SMPL
CCB57 - 1	CCB57	222	0.0354	-7470A	PPB		22 May 2017 10:19:20	U	SMPL
I3242-01DUP - 1	VL-WWTP-003DUP	932	0.2301	-7470A	PPB		22 May 2017 10:21:27	U	SMPL
I3242-01MS - 1	VL-WWTP-003MS	14574	3.9704	-7470A	PPB		22 May 2017 10:23:33	U	SMPL
I3242-01MSD - 1	VL-WWTP-003MSD	15458	4.2128	-7470A	PPB		22 May 2017 10:25:40	U	SMPL
PB99120BL - 1	PBW	0	-0.0255	-7470A	PPB		22 May 2017 10:27:47	U	SMPL
PB99120BS - 1	LCSW	15116	4.1190	-7470A	PPB		22 May 2017 10:29:55	U	SMPL
PB99120TB - 1	PB99060TB	-33	-0.0345	-7470A	PPB		22 May 2017 10:32:02	U	SMPL
I3125-05 - 1	WC-AOC-17-B5-COMP	29	-0.0175	-7470A	PPB		22 May 2017 10:34:11	U	SMPL
I3202-15 - 1	SB-1COMP	39	-0.0148	-7470A	PPB		22 May 2017 10:36:19	U	SMPL
I3202-15DUP - 1	SB-1COMP DUP	47	-0.0126	-7470A	PPB		22 May 2017 10:38:28	U	SMPL
I3202-15MS - 1	SB-1COMPMS	13306	3.6228	-7470A	PPB		22 May 2017 10:40:34	U	SMPL
CCV58 - 1	CCV58	17676	4.8209	-7470A	PPB		22 May 2017 10:42:41	U	SMPL
CCB58 - 1	CCB58	203	0.0302	-7470A	PPB		22 May 2017 10:44:48	U	SMPL
I3202-15MSD - 1	SB-1COMPMSD	13826	3.7653	-7470A	PPB		22 May 2017 10:46:55	U	SMPL
I3202-16 - 1	SB-2COMP	-26	-0.0326	-7470A	PPB		22 May 2017 10:49:02	U	SMPL
I3202-17 - 1	SB-3COMP	41	-0.0142	-7470A	PPB		22 May 2017 10:51:09	U	SMPL
I3202-18 - 1	SB-4COMP	39	-0.0148	-7470A	PPB		22 May 2017 10:53:16	U	SMPL
I3202-19 - 1	SB-5COMP	40	-0.0145	-7470A	PPB		22 May 2017 10:55:23	U	SMPL
I3202-20 - 1	SB-6COMP	41	-0.0142	-7470A	PPB		22 May 2017 10:57:31	U	SMPL
I3202-21 - 1	SB-7COMP	25	-0.0186	-7470A	PPB		22 May 2017 10:59:39	U	SMPL
I3202-22 - 1	SB-8COMP	27	-0.0181	-7470A	PPB		22 May 2017 11:01:47	U	SMPL
I3202-23 - 1	SB-9COMP	42	-0.0139	-7470A	PPB		22 May 2017 11:03:56	U	SMPL
I3202-24 - 1	SB-10COMP	20	-0.0200	-7470A	PPB		22 May 2017 11:06:03	U	SMPL
CCV59 - 1	CCV59	18342	5.0035	-7470A	PPB		22 May 2017 11:08:10	U	SMPL
CCB59 - 1	CCB59	177	0.0231	-7470A	PPB		22 May 2017 11:10:17	U	SMPL
I3202-25 - 1	SB-11COMP	17	-0.0208	-7470A	PPB		22 May 2017 11:12:24	U	SMPL
I3207-04 - 1	TP-11-TP-12	29	-0.0175	-7470A	PPB		22 May 2017 11:14:31	U	SMPL
I3207-08 - 1	TP-9-TP-10	45	-0.0131	-7470A	PPB		22 May 2017 11:16:39	U	SMPL
I3227-03 - 1	RO-11939	43	-0.0137	-7470A	PPB		22 May 2017 11:18:46	U	SMPL
I3228-03 - 1	MH-9	52	-0.0112	-7470A	PPB		22 May 2017 11:20:53	U	SMPL
I3235-03 - 1	TW1-TW2	48	-0.0123	-7470A	PPB		22 May 2017 11:23:01	U	SMPL
I3235-07 - 1	TW1-TW2	62	-0.0085	-7470A	PPB		22 May 2017 11:25:09	U	SMPL
CCV60 - 1	CCV60	17536	4.7825	-7470A	PPB		22 May 2017 11:27:17	U	SMPL
CCB60 - 1	CCB60	200	0.0294	-7470A	PPB		22 May 2017 11:29:26	U	SMPL
I3242-01LX5 - 1		320	0.0623	-7470A	PPB		22 May 2017 11:31:33	U	SMPL
I3242-01A - 1		15350	4.1832	-7470A	PPB		22 May 2017 11:33:40	U	SMPL
I3202-15LX5 - 1		85	-0.0022	-7470A	PPB		22 May 2017 11:35:47	U	SMPL
I3202-15A - 1		14302	3.8958	-7470A	PPB		22 May 2017 11:37:54	U	SMPL
CCV61 - 1	CCV61	18091	4.9347	-7470A	PPB		22 May 2017 11:40:02	U	SMPL
CCB61 - 1	CCB61	167	0.0203	-7470A	PPB		22 May 2017 11:42:09	U	SMPL