



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87604

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/22/2017	12:25
2	CAL2	1	Water			7.01	05/22/2017	12:26
3	CAL3	1	Water			10.02	05/22/2017	12:27
4	ICV	1	Water			7.00	05/22/2017	12:30
5	CCV1	1	Water			2.01	05/22/2017	12:31
6	I3292-01	1	Solid	20.02	20	12.68	05/22/2017	12:32
7	I3292-02	1	Solid	20.01	20	9.24	05/22/2017	12:33
8	I3292-02DUP	1	Solid	20.03	20	9.25	05/22/2017	12:34
9	I3292-03	1	Solid	20.04	20	8.28	05/22/2017	12:35
10	CCV2	1	Water			12.02	05/22/2017	12:36

1087604

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3292

WorkList ID : 98750

Date : 5/22/2017 12:05:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3292-01	Corrosivity	Cool 4 deg C	ATCE02	I31	170505NNWCP-01-01-LCC-PM	05/05/2017	9040B
	Solid	I3292-02	Corrosivity	Cool 4 deg C	ATCE02	I31	170505NNWCP-02-LCC-PPE	05/05/2017	9040B
	Solid	I3292-03	Corrosivity	Cool 4 deg C	ATCE02	I31	170505NNWCP-03-LCC-PS	05/05/2017	9040B

Date/Time 05-22-17 12:05 PM
Received by: SP
Relinquished by: CP

Date/Time 05-22-17 2:30 PM
Received by: CP
Relinquished by: SP