



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87605

Temp.: 23.2 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/22/2017	12:25
2	CAL2	1	Water			7.01	05/22/2017	12:26
3	CAL3	1	Water			10.02	05/22/2017	12:27
4	ICV	1	Water			7.00	05/22/2017	12:30
5	CCV1	1	Water			2.02	05/22/2017	12:40
6	I3292-04	1	Water			4.20	05/22/2017	12:41
7	I3292-04DUP	1	Water			4.21	05/22/2017	12:42
8	CCV2	1	Water			12.01	05/22/2017	12:43

WORKLIST(Hardcopy Internal Chain)

LB87605 2987605

WorkList Name : CORROSIVITYW I3292

WorkList ID : 98751

Date : 5/22/2017 12:06:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3292-04	Corrosivity	Cool 4 deg C	ATCE02	I31	170505NNWCP-04-LLC-WW	05/05/2017	9040C

Date/Time 05-22-17 12:06 PM
 Received by: JP
 Relinquished by: CR

Date/Time 05-22-17 23:03
 Received by: CP
 Relinquished by: JP