



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87606

Temp.: 24.0 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/22/2017	13:30
2	CAL2	1	Water			7.01	05/22/2017	13:31
3	CAL3	1	Water			10.04	05/22/2017	13:32
4	ICV	1	Water			7.00	05/22/2017	13:33
5	CCV1	1	Water			2.01	05/22/2017	13:35
6	I3269-21	1	Solid	20.02	20	7.62	05/22/2017	13:36
7	I3269-21DUP	1	Solid	20.04	20	7.63	05/22/2017	13:37
8	I3269-22	1	Solid	20.01	20	5.92	05/22/2017	13:39
9	I3269-23	1	Solid	20.03	20	7.68	05/22/2017	13:40
10	I3269-24	1	Solid	20.02	20	5.73	05/22/2017	13:41
11	I3269-25	1	Solid	20.00	20	7.04	05/22/2017	13:42
12	I3269-26	1	Solid	20.04	20	6.95	05/22/2017	13:44
13	I3269-27	1	Solid	20.01	20	7.05	05/22/2017	13:45
14	I3269-28	1	Solid	20.02	20	7.72	05/22/2017	13:47
15	I3269-29	1	Solid	20.05	20	7.58	05/22/2017	13:48
16	CCV2	1	Water			12.02	05/22/2017	13:50
17	I3269-30	1	Solid	20.04	20	8.30	05/22/2017	13:51
18	I3271-03	1	Solid	20.00	20	7.01	05/22/2017	13:52
19	I3290-05	1	Solid	20.02	20	5.15	05/22/2017	13:53
20	I3290-10	1	Solid	20.03	20	5.20	05/22/2017	13:55

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			12.02	05/22/2017	14:00

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3270

WorkList ID : 98749

Date : 5/22/2017 12:03:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3269-21	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-01-COMP	05/18/2017	9045C
	Solid	I3269-22	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-02-COMP	05/18/2017	9045C
	Solid	I3269-23	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-05-COMP	05/18/2017	9045C
	Solid	I3269-24	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-07-COMP	05/18/2017	9045C
	Solid	I3269-25	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-09-COMP	05/18/2017	9045C
	Solid	I3269-26	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-13-COMP	05/18/2017	9045C
	Solid	I3269-27	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-03-COMP	05/19/2017	9045C
	Solid	I3269-28	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-06-COMP	05/19/2017	9045C
	Solid	I3269-29	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-12-COMP	05/19/2017	9045C
	Solid	I3269-30	Corrosivity	Cool 4 deg C	LIRO01	I53	SB-14-COMP	05/19/2017	9045C
	Solid	I3271-03	Corrosivity	Cool 4 deg C	LIRO01	I23	SB-17-COMP	05/19/2017	9045C
05/26/2017	Solid	I3290-05	Corrosivity	Cool 4 deg C	PSEG01	G61	TP-3-TP-4	05/19/2017	9045C
05/26/2017	Solid	I3290-10	Corrosivity	Cool 4 deg C	PSEG01	G61	TP-2	05/19/2017	9045C

Date/Time 05.22.17 12.03 PM
 Received by: SR
 Relinquished by: CR

Date/Time 05.22.17 530 PM
 Received by: CP
 Relinquished by: JP