


7470A

LB 87621
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36613



A= 0.0000e+000

B= 2.7180e-004

C= 4.6237e-002

Rho= 0.9999287

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	-49	0.000	-49				
0.2	0.200	0.214	0.014	619	0.0 %	619				
2.5	2.500	2.468	-0.032	8910	0.0 %	8910				
5.0	5.000	4.925	-0.075	17951	0.0 %	17951				
7.5	7.500	7.561	0.061	27649	0.0 %	27649				
10.0	10.000	9.998	-0.002	36613	0.0 %	36613				

LB87621 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 May 2017 08:52:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-49	-	0.0000	7470A	PPB	23 May 2017 09:03:45	S	Std
0.2 - 1	50.2	619	-	0.2000	7470A	PPB	23 May 2017 09:05:53	S	Std
2.5 - 1	52.5	8910	-	2.5000	7470A	PPB	23 May 2017 09:08:02	S	Std
5.0 - 1	55	17951	-	5.0000	7470A	PPB	23 May 2017 09:10:10	S	Std
7.5 - 1	57.5	27649	-	7.5000	7470A	PPB	23 May 2017 09:12:18	S	Std
10.0 - 1	511	36613	-	10.0000	7470A	PPB	23 May 2017 09:14:25	S	Std
ICV68 - 1	ICV68	14260	3.9222	-	7470A	PPB	23 May 2017 09:17:39	U	SMPL
ICB68 - 1	ICB68	-125	0.0123	-	7470A	PPB	23 May 2017 09:19:45	U	SMPL
CCV75 - 1	CCV75	18354	5.0349	-	7470A	PPB	23 May 2017 09:21:52	U	SMPL
CCB75 - 1	CCB75	-130	0.0109	-	7470A	PPB	23 May 2017 09:23:59	U	SMPL
CRA - 1	CRA	614	0.2131	-	7470A	PPB	23 May 2017 09:26:05	U	SMPL
HighStd - 1	HighStd	36249	9.8989	-	7470A	PPB	23 May 2017 09:28:12	U	SMPL
ChkStd - 1	ChkStd	25186	6.8919	-	7470A	PPB	23 May 2017 09:30:19	U	SMPL
PB99185BL - 1	PBW	-166	0.0011	-	7470A	PPB	23 May 2017 09:32:27	U	SMPL
PB99185BS - 1	LCSW	13812	3.8004	-	7470A	PPB	23 May 2017 09:34:34	U	SMPL
PB99185TB - 1	PB99129TB	-141	0.0079	-	7470A	PPB	23 May 2017 09:36:42	U	SMPL
I3243-05 - 1	TP-7-TP-8	-79	0.0248	-	7470A	PPB	23 May 2017 09:38:50	U	SMPL
I3243-10 - 1	TP-5-TP-6	-56	0.0310	-	7470A	PPB	23 May 2017 09:40:58	U	SMPL
I3249-18 - 1	SB-12COMP	-59	0.0302	-	7470A	PPB	23 May 2017 09:43:07	U	SMPL
I3249-20 - 1	SB-13COMP	-58	0.0305	-	7470A	PPB	23 May 2017 09:45:14	U	SMPL
CCV76 - 1	CCV76	18234	5.0023	-	7470A	PPB	23 May 2017 09:47:20	U	SMPL
CCB76 - 1	CCB76	-118	0.0142	-	7470A	PPB	23 May 2017 09:49:27	U	SMPL
I3249-22 - 1	SB-14COMP	-75	0.0259	-	7470A	PPB	23 May 2017 09:51:34	U	SMPL
I3249-24 - 1	SB-19COMP	-67	0.0280	-	7470A	PPB	23 May 2017 09:53:41	U	SMPL
I3249-26 - 1	SB-18COMP	-51	0.0324	-	7470A	PPB	23 May 2017 09:55:48	U	SMPL
I3249-28 - 1	SB-17COMP	-49	0.0329	-	7470A	PPB	23 May 2017 09:57:55	U	SMPL
I3249-30 - 1	SB-16COMP	-49	0.0329	-	7470A	PPB	23 May 2017 10:00:03	U	SMPL
I3249-32 - 1	SB-20COMP	-52	0.0321	-	7470A	PPB	23 May 2017 10:02:10	U	SMPL
I3250-02 - 1	W.P.RD-C.AVE(SP)	-104	0.0180	-	7470A	PPB	23 May 2017 10:04:19	U	SMPL
I3250-02DUP - 1	W.P.RD-C.AVE(SP)DUP	-116	0.0147	-	7470A	PPB	23 May 2017 10:06:27	U	SMPL
I3250-02MS - 1	W.P.RD-C.AVE(SP)MS	13282	3.6563	-	7470A	PPB	23 May 2017 10:08:36	U	SMPL
I3250-02MSD - 1	W.P.RD-C.AVE(SP)MSD	12997	3.5789	-	7470A	PPB	23 May 2017 10:10:43	U	SMPL
CCV77 - 1	CCV77	18244	5.0050	-	7470A	PPB	23 May 2017 10:12:50	U	SMPL
CCB77 - 1	CCB77	-116	0.0147	-	7470A	PPB	23 May 2017 10:14:56	U	SMPL
I3278-03 - 1	RT-3873	-61	0.0297	-	7470A	PPB	23 May 2017 10:17:03	U	SMPL
I3278-06 - 1	RT-2669	-44	0.0343	-	7470A	PPB	23 May 2017 10:19:10	U	SMPL
I3278-09 - 1	RT-4632	-87	0.0226	-	7470A	PPB	23 May 2017 10:21:17	U	SMPL
PB99186BL - 1	PBW	-81	0.0242	-	7470A	PPB	23 May 2017 10:23:24	U	SMPL
PB99186BS - 1	LCSW	14256	3.9211	-	7470A	PPB	23 May 2017 10:25:32	U	SMPL
PB99186TB - 1	PB99145TB	-103	0.0182	-	7470A	PPB	23 May 2017 10:27:39	U	SMPL
I3214-18 - 1	EP2-WC8(0-8)	-39	0.0356	-	7470A	PPB	23 May 2017 10:29:48	U	SMPL
I3214-18DUP - 1	EP2-WC8(0-8)DUP	-41	0.0351	-	7470A	PPB	23 May 2017 10:31:56	U	SMPL
I3214-18MS - 1	EP2-WC8(0-8)MS	14167	3.8969	-	7470A	PPB	23 May 2017 10:34:05	U	SMPL
I3214-18MSD - 1	EP2-WC8(0-8)MSD	13016	3.5840	-	7470A	PPB	23 May 2017 10:36:12	U	SMPL
CCV78 - 1	CCV78	17981	4.9336	-	7470A	PPB	23 May 2017 10:38:19	U	SMPL
CCB78 - 1	CCB78	-112	0.0158	-	7470A	PPB	23 May 2017 10:40:26	U	SMPL
I3214-19 - 1	EP2-WC7(0-6.5)	-47	0.0335	-	7470A	PPB	23 May 2017 10:42:33	U	SMPL
I3214-20 - 1	EP2-WC6(0-6.5)	-49	0.0329	-	7470A	PPB	23 May 2017 10:44:41	U	SMPL
I3214-21 - 1	EP2-WC5(0-10)	-58	0.0305	-	7470A	PPB	23 May 2017 10:46:47	U	SMPL
I3214-22 - 1	EP2-WC5(10-24)	-60	0.0299	-	7470A	PPB	23 May 2017 10:48:54	U	SMPL
I3214-23 - 1	EP2-WC1(6-24)	-68	0.0278	-	7470A	PPB	23 May 2017 10:51:02	U	SMPL
I3215-21 - 1	EP1-WC11(0-2)	-3	0.0454	-	7470A	PPB	23 May 2017 10:53:09	U	SMPL
I3215-22 - 1	EP1-WC11(6-8)	-30	0.0381	-	7470A	PPB	23 May 2017 10:55:18	U	SMPL
I3215-23 - 1	EP1-WC9(0-9.8)	-39	0.0356	-	7470A	PPB	23 May 2017 10:57:25	U	SMPL
I3215-24 - 1	EP1-WC9(9.8-12)	-41	0.0351	-	7470A	PPB	23 May 2017 10:59:34	U	SMPL
I3215-25 - 1	EP1-WC7(0-4)	-59	0.0302	-	7470A	PPB	23 May 2017 11:01:41	U	SMPL
CCV79 - 1	CCV79	18036	4.9485	-	7470A	PPB	23 May 2017 11:03:48	U	SMPL
CCB79 - 1	CCB79	-124	0.0125	-	7470A	PPB	23 May 2017 11:05:55	U	SMPL
I3215-26 - 1	EP1-WC4(0-9.5)	-49	0.0329	-	7470A	PPB	23 May 2017 11:08:02	U	SMPL
I3215-27 - 1	EP1-WC2(0-6)	-31	0.0378	-	7470A	PPB	23 May 2017 11:10:09	U	SMPL
PB99187BL - 1	PBW	-48	0.0332	-	7470A	PPB	23 May 2017 11:12:16	U	SMPL
PB99187BS - 1	LCSW	14217	3.9105	-	7470A	PPB	23 May 2017 11:14:23	U	SMPL
I3245-01 - 1	GMW-7	502	0.1827	-	7470A	PPB	23 May 2017 11:16:31	U	SMPL
I3245-01DUP - 1	GMW-7DUP	520	0.1876	-	7470A	PPB	23 May 2017 11:18:39	U	SMPL
I3245-02 - 1	GMW-X	1232	0.3811	-	7470A	PPB	23 May 2017 11:20:46	U	SMPL
I3245-03 - 1	GMW-10	228	0.1082	-	7470A	PPB	23 May 2017 11:22:54	U	SMPL
I3245-04 - 1	I3245-01MS	14660	4.0309	-	7470A	PPB	23 May 2017 11:25:03	U	SMPL
I3245-05 - 1	I3245-01MSD	14587	4.0111	-	7470A	PPB	23 May 2017 11:27:10	U	SMPL
CCV80 - 1	CCV80	18199	4.9928	-	7470A	PPB	23 May 2017 11:29:18	U	SMPL
CCB80 - 1	CCB80	-117	0.0144	-	7470A	PPB	23 May 2017 11:31:25	U	SMPL
I3245-06 - 1	FB	-118	0.0142	-	7470A	PPB	23 May 2017 11:33:33	U	SMPL
I3245-08 - 1	GMW-7	-53	0.0318	-	7470A	PPB	23 May 2017 11:35:40	U	SMPL
I3245-09 - 1	GMW-X	-60	0.0299	-	7470A	PPB	23 May 2017 11:37:47	U	SMPL
I3245-10 - 1	GMW-10	180	0.0952	-	7470A	PPB	23 May 2017 11:39:54	U	SMPL
I3245-11 - 1	I3245-08MS	13715	3.7740	-	7470A	PPB	23 May 2017 11:42:01	U	SMPL
I3245-12 - 1	I3245-08MSD	14048	3.8645	-	7470A	PPB	23 May 2017 11:44:08	U	SMPL
I3245-13 - 1	FB	-149	0.0057	-	7470A	PPB	23 May 2017 11:46:16	U	SMPL
CCV81 - 1	CCV81	18156	4.9811	-	7470A	PPB	23 May 2017 11:48:24	U	SMPL
CCB81 - 1	CCB81	-83	0.0237	-	7470A	PPB	23 May 2017 11:50:33	U	SMPL
I3250-02LX5 - 1		-41	0.0351	-	7470A	PPB	23 May 2017 11:52:41	U	SMPL
I3250-02A - 1		14383	3.9556	-	7470A	PPB	23 May 2017 11:54:49	U	SMPL

LB87621 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 May 2017 08:52:37

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
I3214-18LX5 - 1		-96	0.0201		-7470A	PPB	23 May 2017 11:56:56	U	SMPL
I3214-18A - 1		14275	3.9262		-7470A	PPB	23 May 2017 11:59:04	U	SMPL
I3245-01LX5 - 1		0	0.0462		-7470A	PPB	23 May 2017 12:01:11	U	SMPL
I3245-01A - 1		15250	4.1913		-7470A	PPB	23 May 2017 12:03:19	U	SMPL
CCV82 - 1	CCV82	17969	4.9303		-7470A	PPB	23 May 2017 12:05:27	U	SMPL
CCB82 - 1	CCB82	-80	0.0245		-7470A	PPB	23 May 2017 12:07:35	U	SMPL