



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB87635

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	05/23/2017	09:05
2	I3235-07		1	>100.0	>212.0	05/23/2017	09:30
3	I3235-07DUP		1	>100.0	>212.0	05/23/2017	09:50
4	I3239-01		1	>100.0	>212.0	05/23/2017	10:15
5	I3239-02		1	>100.0	>212.0	05/23/2017	10:40
6	I3291-01		1	>100.0	>212.0	05/23/2017	11:00
7	I3303-01		1	>100.0	>212.0	05/23/2017	11:15

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 47635

Worklist Name : FP52317

Worklist ID : 98808

Date : 5/23/2017 08:30AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3235-07	Flash Point	Cool 4 deg C	PSEG01	F31	TW1-TW2	05/18/2017	1010A
	Water	I3239-01	Flash Point	Cool 4 deg C	SEAC01	I32	PURGE-WATER-A	05/19/2017	1010A
	Water	I3239-02	Flash Point	Cool 4 deg C	SEAC01	I32	PURGE-WATER-B	05/19/2017	1010A
	Water	I3291-01	Flash Point	Cool 4 deg C	DIST01		2AVE-04	05/19/2017	1010A
	Water	I3303-01	Flash Point	Cool 4 deg C	TULL02	I62	WS-6	05/22/2017	1010A

Date/Time 5/23/17 8:35AM
 Received by: KS
 Relinquished by: CP

Date/Time 5/23/17 11:20AM
 Received by: CP
 Relinquished by: KS