


7471A

LB87643
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43725



A= 0.0000e+000

B= 2.2953e-004

C= -4.7824e-002

Rho= 0.9999577

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	128	0.000	128				
0.2	0.200	0.195	-0.005	1056	0.0 %	1056				
2.5	2.500	2.495	-0.005	11077	0.0 %	11077				
5.0	5.000	5.073	0.073	22310	0.0 %	22310				
7.5	7.500	7.468	-0.032	32743	0.0 %	32743				
10.0	10.000	9.988	-0.012	43725	0.0 %	43725				

LB87643 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 23 May 2017 14:19:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	128	-	0.00007471A	PPB		23 May 2017 14:23:13	S	Std
0.2 - 1	50.2	1056	-	0.20007471A	PPB		23 May 2017 14:25:31	S	Std
2.5 - 1	52.5	11077	-	2.50007471A	PPB		23 May 2017 14:27:49	S	Std
5.0 - 1	55	22310	-	5.00007471A	PPB		23 May 2017 14:30:06	S	Std
7.5 - 1	57.5	32743	-	7.50007471A	PPB		23 May 2017 14:32:23	S	Std
10.0 - 1	510	43725	-	10.00007471A	PPB		23 May 2017 14:34:40	S	Std
ICV69 - 1	ICV69	16947	3.8420	-7471A	PPB		23 May 2017 14:37:31	U	SMPL
ICB69 - 1	ICB69	27	-0.0416	-7471A	PPB		23 May 2017 14:39:47	U	SMPL
CCV85 - 1	CCV85	21557	4.9002	-7471A	PPB		23 May 2017 14:42:03	U	SMPL
CCB85 - 1	CCB85	17	-0.0439	-7471A	PPB		23 May 2017 14:44:19	U	SMPL
CRA - 1	CRA	1024	0.1872	-7471A	PPB		23 May 2017 14:46:34	U	SMPL
HighStd - 1	HighStd	42953	9.8112	-7471A	PPB		23 May 2017 14:48:50	U	SMPL
ChkStd - 1	ChkStd	29852	6.8041	-7471A	PPB		23 May 2017 14:51:06	U	SMPL
PB99191BL - 1	PBS	-126	-0.0767	-7471A	PPB		23 May 2017 14:53:23	U	SMPL
PB99191BS - 1	LCSS	14818	3.3534	-7471A	PPB		23 May 2017 14:55:40	U	SMPL
I3250-01 - 1	W.P.RD-C.AVE(SP)	1878	0.3832	-7471A	PPB		23 May 2017 14:57:57	U	SMPL
I3252-01 - 1	NCC-BUCK-CLAY	280	0.0164	-7471A	PPB		23 May 2017 15:00:14	U	SMPL
I3257-01 - 1	NCC-JM-COMPOST	2732	0.5793	-7471A	PPB		23 May 2017 15:02:32	U	SMPL
CCV86 - 1	CCV86	21278	4.8361	-7471A	PPB		23 May 2017 15:24:12	U	SMPL
CCB86 - 1	CCB86	52	-0.0359	-7471A	PPB		23 May 2017 15:26:27	U	SMPL
I3281-04 - 1	DUP-1	12379	2.7935	-7471A	PPB		23 May 2017 15:33:16	U	SMPL
I3281-05 - 1	R1-TP-2-D1012	3217	0.6906	-7471A	PPB		23 May 2017 15:35:32	U	SMPL
I3281-06 - 1	R1-TP-3-D34	7664	1.7113	-7471A	PPB		23 May 2017 15:37:48	U	SMPL
I3281-07 - 1	I3281-06MS	24272	5.5234	-7471A	PPB		23 May 2017 15:46:42	U	SMPL
I3281-08 - 1	I3281-06MSD	24369	5.5456	-7471A	PPB		23 May 2017 15:48:58	U	SMPL
I3281-09 - 1	DUP-2	16835	3.8163	-7471A	PPB		23 May 2017 15:51:16	U	SMPL
I3281-10 - 1	R1-TP-4-SS2	39230	8.9567	-7471A	PPB		23 May 2017 15:53:33	U	SMPL
I3281-11 - 1	R1-TP5-D34	8105	1.8125	-7471A	PPB		23 May 2017 15:55:49	U	SMPL
CCV87 - 1	CCV87	21602	4.9105	-7471A	PPB		23 May 2017 15:58:05	U	SMPL
CCB87 - 1	CCB87	60	-0.0341	-7471A	PPB		23 May 2017 16:00:21	U	SMPL
I3281-12 - 1	R1-TP6-D1112	1528	0.3029	-7471A	PPB		23 May 2017 16:02:38	U	SMPL
I3281-13 - 1	R1-TP7-D34	8070	1.8045	-7471A	PPB		23 May 2017 16:04:54	U	SMPL
I3281-14 - 1	R1-TP8-SS3	53424	12.2147	-7471A	PPB		23 May 2017 16:07:10	U	SMPL
I3281-15 - 1	R1-TP9-D23	29562	6.7376	-7471A	PPB		23 May 2017 16:09:26	U	SMPL
I3281-16 - 1	R1-TP10-SS4	60060	13.7378	-7471A	PPB		23 May 2017 16:11:43	U	SMPL
I3281-17 - 1	R1-TP11-D78	12519	2.8257	-7471A	PPB		23 May 2017 16:14:00	U	SMPL
I3281-18 - 1	R1-TP12-D1112	7024	1.5644	-7471A	PPB		23 May 2017 16:16:17	U	SMPL
I3281-19 - 1	R1-TP13-SS5	27116	6.1762	-7471A	PPB		23 May 2017 16:18:34	U	SMPL
I3281-20 - 1	R1-TP14-SS6	35615	8.1269	-7471A	PPB		23 May 2017 16:20:52	U	SMPL
I3281-21 - 1	R1-TP15-D46	5959	1.3200	-7471A	PPB		23 May 2017 16:23:08	U	SMPL
CCV88 - 1	CCV88	22169	5.0407	-7471A	PPB		23 May 2017 16:25:24	U	SMPL
CCB88 - 1	CCB88	100	-0.0249	-7471A	PPB		23 May 2017 16:27:41	U	SMPL
PB99192BL - 1	PBS	-12	-0.0506	-7471A	PPB		23 May 2017 16:29:57	U	SMPL
PB99192BS - 1	LVSS	16140	3.6568	-7471A	PPB		23 May 2017 16:32:14	U	SMPL
I3249-02 - 1	SB-12COMP	127	-0.0187	-7471A	PPB		23 May 2017 16:34:30	U	SMPL
I3249-02DUP - 1	SB-12COMPDUP	136	-0.0166	-7471A	PPB		23 May 2017 16:36:46	U	SMPL
I3249-02MS - 1	SB-12COMPMS	15023	3.4004	-7471A	PPB		23 May 2017 16:39:03	U	SMPL
I3249-02MSD - 1	SB-12COMPMSD	14811	3.3518	-7471A	PPB		23 May 2017 16:41:20	U	SMPL
I3249-04 - 1	SB-13COMP	262	0.0123	-7471A	PPB		23 May 2017 16:43:37	U	SMPL
I3249-06 - 1	SB-14COMP	318	0.0252	-7471A	PPB		23 May 2017 16:45:55	U	SMPL
I3249-08 - 1	SB-19COMP	95	-0.0260	-7471A	PPB		23 May 2017 16:48:12	U	SMPL
I3249-10 - 1	SB-18COMP	75	-0.0306	-7471A	PPB		23 May 2017 16:50:28	U	SMPL
CCV89 - 1	CCV89	21912	4.9817	-7471A	PPB		23 May 2017 16:52:45	U	SMPL
CCB89 - 1	CCB89	101	-0.0246	-7471A	PPB		23 May 2017 16:55:01	U	SMPL
I3249-12 - 1	SB-17COMP	96	-0.0258	-7471A	PPB		23 May 2017 16:57:17	U	SMPL
I3249-14 - 1	SB-16COMP	212	0.0008	-7471A	PPB		23 May 2017 16:59:34	U	SMPL
I3249-16 - 1	SB-20COMP	129	-0.0182	-7471A	PPB		23 May 2017 17:01:50	U	SMPL
I3261-01 - 1	NCC-BUCK-3-1	818	0.1399	-7471A	PPB		23 May 2017 17:04:06	U	SMPL
I3264-01 - 1	NCC-BUCK-FILL	178	-0.0070	-7471A	PPB		23 May 2017 17:06:23	U	SMPL
I3278-01 - 1	RT-3873	1326	0.2565	-7471A	PPB		23 May 2017 17:08:39	U	SMPL
I3278-04 - 1	RT-2669	1352	0.2625	-7471A	PPB		23 May 2017 17:10:57	U	SMPL
I3278-07 - 1	RT-4632	1023	0.1870	-7471A	PPB		23 May 2017 17:13:14	U	SMPL
I3290-01 - 1	TP-3-TP-4	445	0.0543	-7471A	PPB		23 May 2017 17:15:32	U	SMPL
I3290-06 - 1	TP-2	158	-0.0116	-7471A	PPB		23 May 2017 17:17:48	U	SMPL
CCV90 - 1	CCV90	21899	4.9787	-7471A	PPB		23 May 2017 17:20:05	U	SMPL
CCB90 - 1	CCB90	99	-0.0251	-7471A	PPB		23 May 2017 17:22:22	U	SMPL
I3300-01 - 1	E2-T9-BASE	4177	0.9109	-7471A	PPB		23 May 2017 17:24:39	U	SMPL
I3300-02 - 1	E2-T10-BASE	21066	4.7875	-7471A	PPB		23 May 2017 17:26:55	U	SMPL
I3300-03 - 1	E2-T11-ESW	10094	2.2691	-7471A	PPB		23 May 2017 17:29:12	U	SMPL
I3300-04 - 1	E2-S11-ESW	18703	4.2451	-7471A	PPB		23 May 2017 17:31:29	U	SMPL
I3300-05 - 1	E2-R11-WSW	4590	1.0057	-7471A	PPB		23 May 2017 17:33:45	U	SMPL
I3281-01 - 1	R1-TP-1-SS-1	5562	1.2288	-7471A	PPB		23 May 2017 17:45:46	U	SMPL
I3281-01DUP - 1	R1-TP-1-SS-1DUP	5473	1.2084	-7471A	PPB		23 May 2017 17:48:02	U	SMPL
I3281-02 - 1	I3281-01MS	22285	5.0673	-7471A	PPB		23 May 2017 17:50:20	U	SMPL
I3281-03 - 1	I3281-01MSD	22146	5.0354	-7471A	PPB		23 May 2017 17:52:38	U	SMPL
CCV91 - 1	CCV91	21393	4.8625	-7471A	PPB		23 May 2017 17:54:55	U	SMPL
CCB91 - 1	CCB91	44	-0.0377	-7471A	PPB		23 May 2017 17:57:12	U	SMPL
I3281-01LX5 - 1		1190	0.2253	-7471A	PPB		23 May 2017 17:59:29	U	SMPL
I3281-01A - 1		22185	5.0443	-7471A	PPB		23 May 2017 18:01:46	U	SMPL
I3249-02LX5 - 1		2	-0.0474	-7471A	PPB		23 May 2017 18:04:03	U	SMPL
I3249-02A - 1		16697	3.7847	-7471A	PPB		23 May 2017 18:06:20	U	SMPL

LB87643 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 23 May 2017 14:19:12					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
I3281-14DLX2 - 1		25541	5.8146		-7471A	PPB	23 May 2017 18:08:37	U	SMPL
I3281-16DLX2 - 1		28930	6.5925		-7471A	PPB	23 May 2017 18:10:54	U	SMPL
CCV92 - 1		22316	5.0744		-7471A	PPB	23 May 2017 18:13:11	U	SMPL
CCB92 - 1		113	-0.0219		-7471A	PPB	23 May 2017 18:15:28	U	SMPL