



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB87655

Temp.: 15.6 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/23/2017	16:35
2	CAL2	1	Water			7.00	05/23/2017	16:36
3	CAL3	1	Water			10.02	05/23/2017	16:37
4	ICV	1	Water			6.99	05/23/2017	16:40
5	CCV1	1	Water			2.01	05/23/2017	16:41
6	I3309-01	1	Water			6.43	05/23/2017	16:42
7	I3309-01DUP	1	Water			6.44	05/23/2017	16:42
8	I3309-02	1	Water			6.81	05/23/2017	16:43
9	I3315-01	1	Water			6.51	05/23/2017	16:45
10	CCV2	1	Water			12.02	05/23/2017	16:46

2B87655

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3309

WorkList ID : 98825

Date : 5/23/2017 4:26:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3309-01-F	pH	Cool 4 deg C	WOOD06	J41	GMW-8	05/23/2017	SM 4500-PH B
	Water	I3309-02-E	pH	Cool 4 deg C	WOOD06	J41	GMW-9	05/23/2017	SM 4500-PH B
	Water	I3315-01-B	pH	Cool 4 deg C	CHEM05	H13	OWS-WATER-INPUT-1	05/23/2017	SM 4500-PH B

Date/Time 05/23/17 4:26 PM
 Received by: 74
 Relinquished by: 74

Date/Time 05/23/17 5:35 PM
 Received by: 74
 Relinquished by: 74