



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87657

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/23/2017	15:25
2	CAL2	1	Water			7.00	05/23/2017	15:26
3	CAL3	1	Water			10.03	05/23/2017	15:27
4	ICV	1	Water			7.00	05/23/2017	15:29
5	CCV1	1	Water			2.01	05/23/2017	15:30
6	I3308-09	1	Solid	20.03	20	7.24	05/23/2017	15:31
7	I3308-09DUP	1	Solid	20.04	20	7.25	05/23/2017	15:32
8	I3308-10	1	Solid	20.01	20	7.27	05/23/2017	15:33
9	I3308-11	1	Solid	20.00	20	6.67	05/23/2017	15:36
10	I3308-12	1	Solid	20.04	20	7.03	05/23/2017	15:37
11	I3310-05	1	Solid	20.01	20	8.30	05/23/2017	15:39
12	CCV2	1	Water			12.01	05/23/2017	15:40

WORKLIST(Hardcopy Internal Chain)

188764 J8 1887657
JP

WorkList ID : 98821

WorkList Name : CORROSIVITY I3310

Date : 5/23/2017 3:06:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3308-09	Corrosivity	Cool 4 deg C	LIRO01	J31	SB-10-COMP	05/22/2017	9045C
	Solid	I3308-10	Corrosivity	Cool 4 deg C	LIRO01	J31	SB-15-COMP	05/22/2017	9045C
	Solid	I3308-11	Corrosivity	Cool 4 deg C	LIRO01	J31	SB-16-COMP	05/22/2017	9045C
	Solid	I3308-12	Corrosivity	Cool 4 deg C	LIRO01	J31	SB-18-COMP	05/22/2017	9045C
05/26/2017	Solid	I3310-05	Corrosivity	Cool 4 deg C	PSEG01	H13	RO-2019	05/23/2017	9045C

Date/Time 05-23-17 21:06 PM
Received by: JP
Relinquished by: _____

Date/Time 05-23-17 530 PM
Received by: _____
Relinquished by: JP