


7470A

LB 87667
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36096



A= 0.0000e+000

B= 2.7856e-004

C= 2.1873e-002

Rho= 0.9998314

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	73	0.000	73				
0.2	0.200	0.230	0.030	748	0.0 %	748				
2.5	2.500	2.505	0.005	8913	0.0 %	8913				
5.0	5.000	4.864	-0.136	17381	0.0 %	17381				
7.5	7.500	7.483	-0.017	26783	0.0 %	26783				
10.0	10.000	10.077	0.077	36096	0.0 %	36096				

LB87667 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 May 2017 09:07:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	73	-	0.0000	7470A	PPB	24 May 2017 09:14:20	S	Std
0.2 - 1	50.2	748	-	0.2000	7470A	PPB	24 May 2017 09:16:28	S	Std
2.5 - 1	52.5	8913	-	2.5000	7470A	PPB	24 May 2017 09:18:37	S	Std
5.0 - 1	55	17381	-	5.0000	7470A	PPB	24 May 2017 09:20:45	S	Std
7.5 - 1	57.5	26783	-	7.5000	7470A	PPB	24 May 2017 09:22:53	S	Std
10.0 - 1	510	36096	-	10.0000	7470A	PPB	24 May 2017 09:25:00	S	Std
ICV70 - 1	ICV70	14240	3.9886	-	7470A	PPB	24 May 2017 09:27:55	U	SMPL
ICB70 - 1	ICB70	-26	0.0146	-	7470A	PPB	24 May 2017 09:30:01	U	SMPL
CCV93 - 1	CCV93	18434	5.1569	-	7470A	PPB	24 May 2017 09:32:08	U	SMPL
CCB93 - 1	CCB93	-56	0.0063	-	7470A	PPB	24 May 2017 09:34:14	U	SMPL
CRA - 1	CRA	721	0.2227	-	7470A	PPB	24 May 2017 09:36:21	U	SMPL
HighStd - 1	HighStd	35488	9.9074	-	7470A	PPB	24 May 2017 09:38:27	U	SMPL
ChkStd - 1	ChkStd	24679	6.8965	-	7470A	PPB	24 May 2017 09:40:34	U	SMPL
PB99225BL - 1	PBW	10	0.0247	-	7470A	PPB	24 May 2017 09:42:42	U	SMPL
PB99225BS - 1	LCSW	14768	4.1357	-	7470A	PPB	24 May 2017 09:44:50	U	SMPL
PB99225TB - 1	PB99181TB	-87	-0.0024	-	7470A	PPB	24 May 2017 09:46:57	U	SMPL
I3269-21 - 1	SB-01-COMP	-22	0.0157	-	7470A	PPB	24 May 2017 09:49:06	U	SMPL
I3269-22 - 1	SB-02-COMP	-29	0.0138	-	7470A	PPB	24 May 2017 09:51:14	U	SMPL
I3269-23 - 1	SB-05-COMP	6	0.0235	-	7470A	PPB	24 May 2017 09:53:23	U	SMPL
I3269-24 - 1	SB-07-COMP	-10	0.0191	-	7470A	PPB	24 May 2017 09:55:30	U	SMPL
CCV94 - 1	CCV94	18417	5.1521	-	7470A	PPB	24 May 2017 09:57:36	U	SMPL
CCB94 - 1	CCB94	-5	0.0205	-	7470A	PPB	24 May 2017 09:59:43	U	SMPL
I3269-25 - 1	SB-09-COMP	-5	0.0205	-	7470A	PPB	24 May 2017 10:01:50	U	SMPL
I3269-26 - 1	SB-13-COMP	-8	0.0196	-	7470A	PPB	24 May 2017 10:03:57	U	SMPL
I3269-27 - 1	SB-03-COMP	-24	0.0152	-	7470A	PPB	24 May 2017 10:06:03	U	SMPL
I3269-28 - 1	SB-06-COMP	-12	0.0185	-	7470A	PPB	24 May 2017 10:08:10	U	SMPL
I3269-29 - 1	SB-12-COMP	-2	0.0213	-	7470A	PPB	24 May 2017 10:10:18	U	SMPL
I3269-30 - 1	SB-14-COMP	15	0.0261	-	7470A	PPB	24 May 2017 10:12:25	U	SMPL
I3271-03 - 1	SB-17-COMP	61	0.0389	-	7470A	PPB	24 May 2017 10:14:34	U	SMPL
I3290-05 - 1	TP-3-TP-4	48	0.0352	-	7470A	PPB	24 May 2017 10:16:42	U	SMPL
I3290-05DUP - 1	TP-3-TP-4DUP	26	0.0291	-	7470A	PPB	24 May 2017 10:18:50	U	SMPL
I3290-05MS - 1	TP-3-TP-4MS	12790	3.5847	-	7470A	PPB	24 May 2017 10:20:57	U	SMPL
CCV95 - 1	CCV95	18340	5.1307	-	7470A	PPB	24 May 2017 10:23:04	U	SMPL
CCB95 - 1	CCB95	-12	0.0185	-	7470A	PPB	24 May 2017 10:25:11	U	SMPL
I3290-05MSD - 1	TP-3-TP-4MSD	12755	3.5749	-	7470A	PPB	24 May 2017 10:27:17	U	SMPL
I3290-10 - 1	TP-2	-53	0.0071	-	7470A	PPB	24 May 2017 10:29:24	U	SMPL
PB99227BL - 1	PBW	-25	0.0149	-	7470A	PPB	24 May 2017 10:31:31	U	SMPL
PB99227BS - 1	LCSW	13589	3.8072	-	7470A	PPB	24 May 2017 10:33:38	U	SMPL
I3239-01 - 1	PURGE-WATER-A	1433724	HIGH	-	7470A	PPB	24 May 2017 10:35:45	X	
I3239-02 - 1	PURGE-WATER-B	306376	HIGH	-	7470A	PPB	24 May 2017 10:37:34	X	
I3282-01 - 1	PZ-02	11024	3.0927	-	7470A	PPB	24 May 2017 10:50:00	U	SMPL
I3282-02 - 1	PZ-15	346	0.1183	-	7470A	PPB	24 May 2017 10:52:11	U	SMPL
I3282-02DUP - 1	PZ-15DUP	334	0.1149	-	7470A	PPB	24 May 2017 11:02:34	U	SMPL
I3282-02MS - 1	PZ-15MS	16086	4.5028	-	7470A	PPB	24 May 2017 11:04:40	U	SMPL
CCV96 - 1	CCV96	18318	5.1245	-	7470A	PPB	24 May 2017 11:06:47	U	SMPL
CCB96 - 1	CCB96	-272	-0.0539	-	7470A	PPB	24 May 2017 11:08:54	U	SMPL
I3282-02MSD - 1	PZ-15MSD	15910	4.4538	-	7470A	PPB	24 May 2017 11:11:01	U	SMPL
I3282-03 - 1	PZ-01	21028	5.8794	-	7470A	PPB	24 May 2017 11:13:08	U	SMPL
I3282-04 - 1	GMW-1	987	0.2968	-	7470A	PPB	24 May 2017 11:15:15	U	SMPL
I3282-05 - 1	GMW-11	8665	2.4356	-	7470A	PPB	24 May 2017 11:17:23	U	SMPL
I3282-06 - 1	GMW-2	6558	1.8487	-	7470A	PPB	24 May 2017 11:19:30	U	SMPL
I3282-07 - 1	GMW-5	965	0.2907	-	7470A	PPB	24 May 2017 11:21:38	U	SMPL
I3282-09 - 1	PZ-02	-102	-0.0065	-	7470A	PPB	24 May 2017 11:23:46	U	SMPL
I3282-10 - 1	PZ-15	-90	-0.0032	-	7470A	PPB	24 May 2017 11:25:54	U	SMPL
I3282-11 - 1	PZ-01	-37	0.0116	-	7470A	PPB	24 May 2017 11:28:03	U	SMPL
I3282-12 - 1	GMW-1	-58	0.0057	-	7470A	PPB	24 May 2017 11:30:10	U	SMPL
CCV97 - 1	CCV97	17339	4.8518	-	7470A	PPB	24 May 2017 11:32:18	U	SMPL
CCB97 - 1	CCB97	-76	0.0007	-	7470A	PPB	24 May 2017 11:34:25	U	SMPL
I3282-13 - 1	GMW-11	-54	0.0068	-	7470A	PPB	24 May 2017 11:36:32	U	SMPL
I3282-14 - 1	GMW-2	-64	0.0040	-	7470A	PPB	24 May 2017 11:38:39	U	SMPL
I3282-15 - 1	GMW-5	-56	0.0063	-	7470A	PPB	24 May 2017 11:40:46	U	SMPL
I3239-01DLX10 - 1		991532	HIGH	-	7470A	PPB	24 May 2017 11:42:54	X	
I3239-02DLX10 - 1		35892	10.0200	-	7470A	PPB	24 May 2017 11:44:43	U	SMPL
I3290-05LX5 - 1		-212	-0.0372	-	7470A	PPB	24 May 2017 11:47:11	U	SMPL
I3290-05A - 1		14404	4.0343	-	7470A	PPB	24 May 2017 11:49:20	U	SMPL
I3282-02LX5 - 1		18	0.0269	-	7470A	PPB	24 May 2017 11:54:15	U	SMPL
I3282-02A - 1		16333	4.5716	-	7470A	PPB	24 May 2017 11:56:23	U	SMPL
CCV98 - 1	CCV98	17563	4.9142	-	7470A	PPB	24 May 2017 11:58:30	U	SMPL
CCB98 - 1	CCB98	-42	0.0102	-	7470A	PPB	24 May 2017 12:00:38	U	SMPL
I3239-02DL2X50 - 1		7419	2.0885	-	7470A	PPB	24 May 2017 12:07:48	U	SMPL
I3239-01DL2X2000 - 1		16288	4.5591	-	7470A	PPB	24 May 2017 12:11:41	U	SMPL
CCV99 - 1	CCV99	18296	5.1184	-	7470A	PPB	24 May 2017 12:13:48	U	SMPL
CCB99 - 1	CCB99	-22	0.0157	-	7470A	PPB	24 May 2017 12:15:56	U	SMPL