


ISM02.4

LB 87674
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41359



A= 0.0000e+000

B= 2.4214e-004

C= -2.2005e-002

Rho= 0.9999605

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.028	0.028	207	0.000	207					-
0.2	0.200	0.171	-0.029	797	0.0 %	797					-15
2.5	2.500	2.524	0.024	10516	0.0 %	10516					1
5.0	5.000	4.947	-0.053	20521	0.0 %	20521					-1
7.5	7.500	7.537	0.037	31217	0.0 %	31217					0
10.0	10.000	9.993	-0.007	41359	0.0 %	41359					0

NB
5/24/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	207	-		0 ISM02.4	PPB	5/24/2017 12:42	Std
	0.2 S0.2	797	-		0.2 ISM02.4	PPB	5/24/2017 12:44	Std
	2.5 S2.5	10516	-		2.5 ISM02.4	PPB	5/24/2017 12:46	Std
	5 S5	20521	-		5 ISM02.4	PPB	5/24/2017 12:49	Std
	7.5 S7.5	31217	-		7.5 ISM02.4	PPB	5/24/2017 12:51	Std
	10 S10	41359	-		10 ISM02.4	PPB	5/24/2017 12:53	Std
ICV	ICV	16384	3.9452 -		ISM02.4	PPB	5/24/2017 12:56	SMPL
ICB	ICB	163	0.0175 -		ISM02.4	PPB	5/24/2017 12:58	SMPL
CCV001	CCV001	20520	4.9467 -		ISM02.4	PPB	5/24/2017 13:00	SMPL
CCB001	CCB001	157	0.016 -		ISM02.4	PPB	5/24/2017 13:03	SMPL
PB99220BL	PBW003	53	-0.0092 -		ISM02.4	PPB	5/24/2017 13:05	SMPL
I3178-01	MBDCC6	495	0.0979 -		ISM02.4	PPB	5/24/2017 13:12	SMPL
I3178-02	MBDCC6D	454	0.0879 -		ISM02.4	PPB	5/24/2017 13:14	SMPL
I3178-03	MBDCC6S	4697	1.1153 -		ISM02.4	PPB	5/24/2017 13:17	SMPL
I3178-04	MBDCK8	258	0.0405 -		ISM02.4	PPB	5/24/2017 13:19	SMPL
PB99221BL	PBW003	72	-0.0046 -		ISM02.4	PPB	5/24/2017 13:21	SMPL
I3179-01	MBDCM6	177	0.0209 -		ISM02.4	PPB	5/24/2017 13:24	SMPL
I3179-02	MBDCM6D	273	0.0441 -		ISM02.4	PPB	5/24/2017 13:26	SMPL
I3179-03	MBDCM6S	4327	1.0257 -		ISM02.4	PPB	5/24/2017 13:28	SMPL
PB99222BL	PBW003	59	-0.0077 -		ISM02.4	PPB	5/24/2017 13:30	SMPL
I3289-01	MBD7K3	362	0.0656 -		ISM02.4	PPB	5/24/2017 13:33	SMPL
I3289-02	MBD7K3D	371	0.0678 -		ISM02.4	PPB	5/24/2017 13:35	SMPL
I3289-03	MBD7K3S	4498	1.0671 -		ISM02.4	PPB	5/24/2017 13:37	SMPL
CCV002	CCV002	20489	4.9392 -		ISM02.4	PPB	5/24/2017 13:39	SMPL
CCB002	CCB002	164	0.0177 -		ISM02.4	PPB	5/24/2017 13:42	SMPL