



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 05/25/2017DATE & TIME IN: 05/24/2017 16:30OVEN TEMP: 98 °CRun Number: LB87676DATE & TIME OUT: 05/25/2017 12:40OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB87676BL	LB87676BL	89.7115	89.7115	100	89.7115	89.7115	89.7115	0	0
2	I3316-01	WS-3	84.6348	84.6348	100	84.8403	84.8403	84.8403	0.2055	2055
3	I3316-01DUP	WS-3DUP	86.3023	86.3023	100	86.5082	86.5082	86.5082	0.2059	2059

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB 87676

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I3316

WorkList ID : 98869

Date : 5/24/2017 1:40:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3316-01-F	TS	Cool 4 deg C	TULL02	I32	WS-3	05/23/2017	SM2540 B

Date/Time 052417-11h07
Received by: JP
Relinquished by: JP

Date/Time 05-24-17 5:00 B
Received by: JP
Relinquished by: JP