


7471A

LB87683
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43870



A= 0.0000e+000

B= 2.2903e-004

C= 1.6446e-002

Rho= 0.9999324

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	9	0.000	9				
0.2	0.200	0.210	0.010	844	0.0 %	844				
2.5	2.500	2.496	-0.004	10828	0.0 %	10828				
5.0	5.000	4.992	-0.008	21723	0.0 %	21723				
7.5	7.500	7.420	-0.080	32325	0.0 %	32325				
10.0	10.000	10.064	0.064	43870	0.0 %	43870				

LB87683 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 24 May 2017 16:24:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.00007471A	PPB		24 May 2017 16:27:25	S	Std
0.2 - 1	50.2	844	-	0.20007471A	PPB		24 May 2017 16:29:42	S	Std
2.5 - 1	52.5	10828	-	2.50007471A	PPB		24 May 2017 16:32:00	S	Std
5.0 - 1	55	21723	-	5.00007471A	PPB		24 May 2017 16:34:17	S	Std
7.5 - 1	57.5	32325	-	7.50007471A	PPB		24 May 2017 16:36:34	S	Std
10.0 - 1	510	43870	-	10.00007471A	PPB		24 May 2017 16:38:51	S	Std
ICV72 - 1	ICV72	16608	3.8202	-7471A	PPB		24 May 2017 16:41:43	U	SMPL
ICB72 - 1	ICB72	-64	0.0018	-7471A	PPB		24 May 2017 16:43:58	U	SMPL
CCV06 - 1	CCV06	21600	4.9635	-7471A	PPB		24 May 2017 16:46:14	U	SMPL
CCB06 - 1	CCB06	-74	-0.0005	-7471A	PPB		24 May 2017 16:48:30	U	SMPL
CRA - 1	CRA	830	0.2065	-7471A	PPB		24 May 2017 16:50:45	U	SMPL
HighStd - 1	HighStd	43278	9.9284	-7471A	PPB		24 May 2017 16:53:01	U	SMPL
ChkStd - 1	ChkStd	29863	6.8559	-7471A	PPB		24 May 2017 16:55:17	U	SMPL
PB99231BL - 1	PBS	-96	-0.0055	-7471A	PPB		24 May 2017 16:57:34	U	SMPL
PB99231BS - 1	LCSS	14775	3.4003	-7471A	PPB		24 May 2017 16:59:51	U	SMPL
I3302-01 - 1	NM2-COMP-0-6	16581	3.8140	-7471A	PPB		24 May 2017 17:02:08	U	SMPL
I3302-02 - 1	NM2-COMP-6-18	79669	18.2629	-7471A	PPB		24 May 2017 17:04:26	U	SMPL
I3302-03 - 1	NM16-COMP-0-6	9409	2.1714	-7471A	PPB		24 May 2017 17:06:43	U	SMPL
I3302-04 - 1	NM16-COMP-6-18	8891	2.0527	-7471A	PPB		24 May 2017 17:09:02	U	SMPL
I3302-05 - 1	NM17-COMP-0-6	11015	2.5392	-7471A	PPB		24 May 2017 17:11:18	U	SMPL
CCV07 - 1	CCV07	21629	4.9701	-7471A	PPB		24 May 2017 17:13:34	U	SMPL
CCB07 - 1	CCB07	-58	0.0032	-7471A	PPB		24 May 2017 17:15:50	U	SMPL
I3302-06 - 1	NM17-COMP-6-18	14711	3.3857	-7471A	PPB		24 May 2017 17:18:06	U	SMPL
I3302-07 - 1	NM12-COMP-0-6	7367	1.7037	-7471A	PPB		24 May 2017 17:20:22	U	SMPL
I3302-08 - 1	NM12-COMP-6-18	14642	3.3699	-7471A	PPB		24 May 2017 17:22:38	U	SMPL
I3302-09 - 1	NM15-COMP-0-6	9879	2.2790	-7471A	PPB		24 May 2017 17:24:54	U	SMPL
I3302-10 - 1	NM15-COMP-6-18	9429	2.1760	-7471A	PPB		24 May 2017 17:27:11	U	SMPL
I3305-01 - 1	DSK-SB1(5-10)	1620	0.3875	-7471A	PPB		24 May 2017 17:36:45	U	SMPL
I3305-01DUP - 1	DSK-SB1(5-10)DUP	1580	0.3783	-7471A	PPB		24 May 2017 17:39:01	U	SMPL
I3305-02 - 1	I3305-01MS	16303	3.7503	-7471A	PPB		24 May 2017 17:41:18	U	SMPL
I3305-03 - 1	I3305-01MSD	15885	3.6546	-7471A	PPB		24 May 2017 17:43:36	U	SMPL
I3305-04 - 1	DSK-SB2(10-11.8)	1294	0.3128	-7471A	PPB		24 May 2017 17:45:52	U	SMPL
CCV08 - 1	CCV08	21808	5.0111	-7471A	PPB		24 May 2017 17:48:09	U	SMPL
CCB08 - 1	CCB08	-55	0.0038	-7471A	PPB		24 May 2017 17:50:25	U	SMPL
I3305-05 - 1	DSK-SB3(5-10)	1311	0.3167	-7471A	PPB		24 May 2017 17:52:41	U	SMPL
I3310-01 - 1	RO-2019	106030	24.3004	-7471A	PPB		24 May 2017 17:54:57	U	SMPL
I3305-01LX5 - 1		164	0.0540	-7471A	PPB		24 May 2017 17:57:13	U	SMPL
I3305-01A - 1		17851	4.1048	-7471A	PPB		24 May 2017 17:59:31	U	SMPL
I3302-02DLX5 - 1		14675	3.3774	-7471A	PPB		24 May 2017 18:04:06	U	SMPL
I3310-01DLX5 - 1		20172	4.6364	-7471A	PPB		24 May 2017 18:06:22	U	SMPL
CCV09 - 1	CCV09	21923	5.0374	-7471A	PPB		24 May 2017 18:08:40	U	SMPL
CCB09 - 1	CCB09	-47	0.0057	-7471A	PPB		24 May 2017 18:10:57	U	SMPL