



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB87687

Temp.: 23.6 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	05/24/2017	17:15
2	CAL2	1	Water			7.01	05/24/2017	17:16
3	CAL3	1	Water			10.02	05/24/2017	17:17
4	ICV	1	Water			7.00	05/24/2017	17:20
5	CCV1	1	Water			2.02	05/24/2017	17:21
6	I3331-16	1	Solid	20.02	20	6.00	05/24/2017	17:22
7	I3331-16DUP	1	Solid	20.04	20	6.01	05/24/2017	17:23
8	I3331-17	1	Solid	20.05	20	6.32	05/24/2017	17:25
9	I3331-18	1	Solid	20.01	20	6.09	05/24/2017	17:26
10	I3331-19	1	Solid	20.02	20	8.56	05/24/2017	17:30
11	I3331-20	1	Solid	20.01	20	8.13	05/24/2017	17:31
12	I3332-13	1	Solid	20.03	20	8.50	05/24/2017	17:32
13	I3332-14	1	Solid	20.05	20	7.03	05/24/2017	17:33
14	I3332-15	1	Solid	20.02	20	7.98	05/24/2017	17:34
15	I3332-16	1	Solid	20.03	20	8.03	05/24/2017	17:35
16	CCV2	1	Water			12.01	05/24/2017	17:36

WORKLIST(Hardcopy Internal Chain)

1887687

WorkList Name : CORROSIONITY I3331

WorkList ID : 98878

Date : 5/24/2017 5:09:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3331-16	Corrosivity	Cool 4 deg C	TULL02	I61	SB-22COMP	05/24/2017	9045C
	Solid	I3331-17	Corrosivity	Cool 4 deg C	TULL02	I61	SB-23COMP	05/24/2017	9045C
	Solid	I3331-18	Corrosivity	Cool 4 deg C	TULL02	I61	SB-21COMP	05/24/2017	9045C
	Solid	I3331-19	Corrosivity	Cool 4 deg C	TULL02	I61	SB-24COMP	05/24/2017	9045C
	Solid	I3331-20	Corrosivity	Cool 4 deg C	TULL02	I61	SB-26COMP	05/24/2017	9045C
	Solid	I3332-13	Corrosivity	Cool 4 deg C	TULL02	I61	SB-25COMP	05/24/2017	9045C
	Solid	I3332-14	Corrosivity	Cool 4 deg C	TULL02	I61	SB-27COMP	05/24/2017	9045C
	Solid	I3332-15	Corrosivity	Cool 4 deg C	TULL02	I61	SB-28COMP	05/24/2017	9045C
	Solid	I3332-16	Corrosivity	Cool 4 deg C	TULL02	I61	SB-15COMP	05/24/2017	9045C

Date/Time 05/24/17 5:00 PM
 Received by: JP
 Relinquished by: CP

Date/Time 05/24/17 8:30 AM
 Received by: CP
 Relinquished by: JP