


ISM02.4

LB 87689  
INSTRUMENT ID: CV1

Linear

A= 0.0000e+000

B= 2.4053e-004

C= -1.8246e-002

Rho= 0.9999927

Accept=Accepted

Slope  
Y-intercept

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.00   | 0.000  | 0.019  | 0.019  | 156   | 0.000      | 156   |       |       |       |       | -   |
| 0.2    | 0.200  | 0.178  | -0.022 | 815   | 0.0 %      | 815   |       |       |       |       | -11 |
| 2.5    | 2.500  | 2.512  | 0.012  | 10521 | 0.0 %      | 10521 |       |       |       |       | 0   |
| 5.0    | 5.000  | 4.989  | -0.011 | 20816 | 0.0 %      | 20816 |       |       |       |       | 0   |
| 7.5    | 7.500  | 7.496  | -0.004 | 31242 | 0.0 %      | 31242 |       |       |       |       | 0   |
| 10.0   | 10.000 | 10.006 | 0.006  | 41675 | 0.0 %      | 41675 |       |       |       |       | 0   |

NB  
5/24/17

## INSTRUMENT ID : CV1

| Sample ID | Extended ID | $\mu$ Abs. | Conc.     | Std Conc | Method      | Units | Date            | Type |
|-----------|-------------|------------|-----------|----------|-------------|-------|-----------------|------|
|           | 0 S0        | 156        | -         |          | 0 ISM02.4   | PPB   | 5/24/2017 19:31 | Std  |
|           | 0.2 S0.2    | 815        | -         |          | 0.2 ISM02.4 | PPB   | 5/24/2017 19:34 | Std  |
|           | 2.5 S2.5    | 10521      | -         |          | 2.5 ISM02.4 | PPB   | 5/24/2017 19:36 | Std  |
|           | 5 S5        | 20816      | -         |          | 5 ISM02.4   | PPB   | 5/24/2017 19:38 | Std  |
|           | 7.5 S7.5    | 31242      | -         |          | 7.5 ISM02.4 | PPB   | 5/24/2017 19:40 | Std  |
|           | 10 S10      | 41675      | -         |          | 10 ISM02.4  | PPB   | 5/24/2017 19:43 | Std  |
| ICV       | ICV         | 15782      | 3.7778 -  |          | ISM02.4     | PPB   | 5/24/2017 19:45 | SMPL |
| ICB       | ICB         | 74         | -0.0004 - |          | ISM02.4     | PPB   | 5/24/2017 19:48 | SMPL |
| CCV010    | CCV010      | 20852      | 4.9972 -  |          | ISM02.4     | PPB   | 5/24/2017 19:50 | SMPL |
| CCB010    | CCB010      | 76         | 0 -       |          | ISM02.4     | PPB   | 5/24/2017 19:52 | SMPL |
| PB99252BL | PBW003      | 147        | 0.0171 -  |          | ISM02.4     | PPB   | 5/24/2017 19:54 | SMPL |
| I3213-07  | MX2995      | 37211      | 8.932 -   |          | ISM02.4     | PPB   | 5/24/2017 19:57 | SMPL |
| I3213-08  | MX2995D     | 36213      | 8.692 -   |          | ISM02.4     | PPB   | 5/24/2017 19:59 | SMPL |
| I3213-09  | MX2995S     | 40604      | 9.7481 -  |          | ISM02.4     | PPB   | 5/24/2017 20:01 | SMPL |
| CCV011    | CCV011      | 20477      | 4.907 -   |          | ISM02.4     | PPB   | 5/24/2017 20:03 | SMPL |
| CCB011    | CCB011      | 9          | -0.0161 - |          | ISM02.4     | PPB   | 5/24/2017 20:06 | SMPL |