

7470A


LB87691

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35320



A= 0.0000e+000

B= 2.8317e-004

C= 2.0473e-002

Rho= 0.9999480

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	89	0.000	89				
0.2	0.200	0.191	-0.009	603	0.0 %	603				
2.5	2.500	2.494	-0.006	8734	0.0 %	8734				
5.0	5.000	4.927	-0.073	17326	0.0 %	17326				
7.5	7.500	7.521	0.021	26487	0.0 %	26487				
10.0	10.000	10.022	0.022	35320	0.0 %	35320				

LB87691 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 May 2017 08:29:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	89	-	0.0000	7470A	PPB	25 May 2017 08:40:15	S	Std
0.2 - 1	S0.2	603	-	0.2000	7470A	PPB	25 May 2017 08:42:23	S	Std
2.5 - 1	S2.5	8734	-	2.5000	7470A	PPB	25 May 2017 08:44:31	S	Std
5.0 - 1	S5	17326	-	5.0000	7470A	PPB	25 May 2017 08:46:40	S	Std
7.5 - 1	S7.5	26487	-	7.5000	7470A	PPB	25 May 2017 08:53:19	S	Std
10.0 - 1	S10	35320	-	10.0000	7470A	PPB	25 May 2017 08:55:26	S	Std
ICV73 - 1	ICV73	13596	3.8704	-	7470A	PPB	25 May 2017 08:57:57	U	SMPL
ICB73 - 1	ICB73	-3	0.0196	-	7470A	PPB	25 May 2017 09:00:02	U	SMPL
CCV12 - 1	CCV12	18457	5.2469	-	7470A	PPB	25 May 2017 09:02:09	U	SMPL
CCB12 - 1	CCB12	-2	0.0199	-	7470A	PPB	25 May 2017 09:04:16	U	SMPL
CRA - 1	CRA	563	0.1799	-	7470A	PPB	25 May 2017 09:06:22	U	SMPL
HighStd - 1	HighStd	34772	9.8668	-	7470A	PPB	25 May 2017 09:08:29	U	SMPL
ChkStd - 1	ChkStd	24438	6.9406	-	7470A	PPB	25 May 2017 09:10:35	U	SMPL
PB99271BL - 1	PBW	-30	0.0120	-	7470A	PPB	25 May 2017 09:12:43	U	SMPL
PB99271BS - 1	LCSW	14097	4.0123	-	7470A	PPB	25 May 2017 09:14:51	U	SMPL
I3309-01 - 1	GMW-8	153	0.0638	-	7470A	PPB	25 May 2017 09:21:32	U	SMPL
I3309-01DUP - 1	GMW-8DUP	130	0.0573	-	7470A	PPB	25 May 2017 09:23:40	U	SMPL
I3309-01MS - 1	GMW-8MS	14097	4.0123	-	7470A	PPB	25 May 2017 09:25:48	U	SMPL
I3309-01MSD - 1	GMW-8MSD	14698	4.1825	-	7470A	PPB	25 May 2017 09:27:57	U	SMPL
I3309-02 - 1	GMW-9	3573	1.0322	-	7470A	PPB	25 May 2017 09:30:04	U	SMPL
CCV13 - 1	CCV13	18603	5.2883	-	7470A	PPB	25 May 2017 09:32:10	U	SMPL
CCB13 - 1	CCB13	27	0.0281	-	7470A	PPB	25 May 2017 09:34:17	U	SMPL
I3309-03 - 1	GMW-8	2	0.0210	-	7470A	PPB	25 May 2017 09:36:24	U	SMPL
I3309-04 - 1	GMW-9	-93	-0.0059	-	7470A	PPB	25 May 2017 09:38:31	U	SMPL
PB99272BL - 1	PBW	113	0.0525	-	7470A	PPB	25 May 2017 09:40:38	U	SMPL
PB99272BS - 1	LCSW	14181	4.0361	-	7470A	PPB	25 May 2017 09:42:44	U	SMPL
PB99272TB - 1	PB99224TB	40	0.0318	-	7470A	PPB	25 May 2017 09:44:52	U	SMPL
I3308-09 - 1	SB-10-COMP	81	0.0434	-	7470A	PPB	25 May 2017 09:47:00	U	SMPL
I3308-10 - 1	SB-15-COMP	81	0.0434	-	7470A	PPB	25 May 2017 09:49:08	U	SMPL
I3308-11 - 1	SB-16-COMP	49	0.0343	-	7470A	PPB	25 May 2017 09:51:16	U	SMPL
I3308-12 - 1	SB-18-COMP	17	0.0253	-	7470A	PPB	25 May 2017 09:53:25	U	SMPL
I3310-05 - 1	RO-2019	7	0.0225	-	7470A	PPB	25 May 2017 09:55:32	U	SMPL
CCV14 - 1	CCV14	18439	5.2418	-	7470A	PPB	25 May 2017 09:57:39	U	SMPL
CCB14 - 1	CCB14	-4	0.0193	-	7470A	PPB	25 May 2017 09:59:45	U	SMPL
I3310-05DUP - 1	RO-2019DUP	6	0.0222	-	7470A	PPB	25 May 2017 10:01:52	U	SMPL
I3310-05MS - 1	RO-2019MS	13409	3.8175	-	7470A	PPB	25 May 2017 10:04:00	U	SMPL
I3310-05MSD - 1	RO-2019MSD	13321	3.7926	-	7470A	PPB	25 May 2017 10:06:06	U	SMPL
I3309-01LX5 - 1		445	0.1465	-	7470A	PPB	25 May 2017 10:08:13	U	SMPL
I3309-01A - 1		14642	4.1666	-	7470A	PPB	25 May 2017 10:10:20	U	SMPL
I3310-05LX5 - 1		51	0.0349	-	7470A	PPB	25 May 2017 10:12:28	U	SMPL
I3310-05A - 1		14218	4.0466	-	7470A	PPB	25 May 2017 10:14:36	U	SMPL
CCV15 - 1	CCV15	18602	5.2880	-	7470A	PPB	25 May 2017 10:16:44	U	SMPL
CCB15 - 1	CCB15	16	0.0250	-	7470A	PPB	25 May 2017 10:18:52	U	SMPL